

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112624\
 Data File : BN035321.D
 Acq On : 26 Nov 2024 18:55
 Operator : RC/JU
 Sample : P5016-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E1PJ2

Quant Time: Nov 26 22:14:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 26 06:09:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.314	152	3741	0.400	ng/ul	0.00
4) Naphthalene-d8	10.059	136	10114	0.400	ng/ul	0.00
9) Acenaphthene-d10	13.972	164	6346	0.400	ng/ul	0.00
13) Phenanthrene-d10	0.000	188	0	0.000	ng/ul	-16.74
17) Chrysene-d12	20.976	240	7179	0.400	ng/ul	# 0.00
23) Perylene-d12	22.943	264	614739	0.400	ng/ul	#-0.13
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.972	96	16981	4.966	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.659	152	4631	0.312	ng/ul	0.00
18) Fluoranthene-d10	18.786	212	9822	0.496	ng/ul	0.00
Target Compounds						
					Qvalue	
20) Pyrene	19.153	202	1705	0.066	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN11624\
Data File : BN035321.D
Acq On : 26 Nov 2024 18:55
Operator : RC/JU
Sample : P5016-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E1PJ2

Quant Time: Nov 26 22:14:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN11624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 26 06:09:15 2024
Response via : Initial Calibration





#20

Pyrene

Concen: 0.066 ng/ul

RT: 19.153 min Scan# 3

Delta R.T. -0.032 min

Lab File: BN035321.D


Acq: 26 Nov 2024 18:55

Instrument :

BNA_N

ClientSampleId :

E1PJ2



Tgt Ion:202 Resp: 1705

Ion Ratio Lower Upper

202 100

101 147.4 6.2 9.4#

100 1086.1 5.1 7.7#

