

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112624\
 Data File : BN035326.D
 Acq On : 26 Nov 2024 21:55
 Operator : RC/JU
 Sample : P4996-06
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHF73

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/27/2024
 Supervised By :mohammad ahmed 11/28/2024

Quant Time: Nov 26 22:57:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN112624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 26 06:09:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.310	152	3355	0.400	ng/ul	0.00
4) Naphthalene-d8	10.059	136	9317	0.400	ng/ul	0.00
9) Acenaphthene-d10	13.972	164	7286	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.734	188	12884	0.400	ng/ul	# 0.00
17) Chrysene-d12	20.976	240	8547	0.400	ng/ul	# 0.00
23) Perylene-d12	23.068	264	7147m	0.400	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.972	96	14041	4.579	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.659	152	980	0.072	ng/ul	0.00
18) Fluoranthene-d10	18.786	212	11407	0.484	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112624\
Data File : BN035326.D
Acq On : 26 Nov 2024 21:55
Operator : RC/JU
Sample : P4996-06
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHF73

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/27/2024
Supervised By :mohammad ahmed 11/28/2024

Quant Time: Nov 26 22:57:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN112624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 26 06:09:15 2024
Response via : Initial Calibration

