


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112624\
Data File : BN035301.D
Acq On : 26 Nov 2024 04:29
Operator : RC/JU
Sample : DFTPP246
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP246

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Nov 26 06:09:15 2024



AutoFind: Scans 2437, 2438, 2439; Background Corrected with Scan 2431

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.2	37021	PASS
68	69	0.00	2	0.7	351	PASS
69	198	0.00	100	31.4	50115	PASS
70	69	0.00	2	0.5	233	PASS
127	198	10	80	41.4	65933	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	159376	PASS
199	198	5	9	7.0	11202	PASS
275	198	10	60	25.8	41192	PASS
365	198	1	100	3.9	6280	PASS
441	198	0.01	100	10.1	16099	PASS
442	198	50	100	60.2	95920	PASS
443	442	15	24	20.7	19869	PASS

m/z	Abundance
39.90	324.0
44.00	1098.0
73.00	163.0
95.90	199.0
207.00	1036.0
207.90	359.0
281.00	230.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

m/z	Abundance
37.90	291.0
39.00	1877.0
40.00	475.0
44.00	869.0
49.00	398.0
50.10	7526.0
51.00	29256.0
52.00	1614.0
54.90	255.0
56.00	874.0
57.00	2045.0
60.90	388.0
62.00	597.0
63.00	1299.0
64.00	251.0
65.00	685.0
69.00	39136.0
73.00	598.0
74.00	4555.0
75.00	6985.0
76.10	2070.0
77.00	43264.0
78.00	3062.0
79.00	3361.0
80.00	2408.0
81.00	3282.0
82.00	900.0
83.00	876.0
85.00	648.0
85.90	1148.0
87.00	404.0
88.00	150.0
90.90	676.0
92.00	958.0
93.00	5837.0
93.90	383.0
96.00	320.0
97.00	164.0
98.00	4670.0
99.00	3781.0
100.00	377.0
101.00	1952.0
103.00	593.0
103.90	1159.0
105.00	1056.0
106.00	437.0
107.00	14368.0
108.00	2165.0
110.00	25944.0
111.00	3806.0
112.00	588.0
112.90	155.0
117.00	13848.0
118.00	1218.0
119.90	173.0
121.90	1128.0
122.90	1447.0
124.00	949.0
124.90	650.0
127.00	49280.0
128.00	3964.0
129.00	19768.0
130.00	1934.0
131.00	324.0
132.00	168.0
132.70	176.0
133.90	627.0
135.00	1700.0
135.90	786.0
137.00	1111.0
137.90	259.0
140.00	233.0
141.00	2792.0
142.00	846.0
143.00	681.0
144.00	154.0
144.80	151.0
145.90	537.0
146.90	1576.0
147.90	3329.0
148.90	895.0
149.90	210.0
151.00	346.0
151.90	250.0
153.00	835.0
153.90	630.0
155.00	1574.0
156.00	2297.0
157.10	391.0
157.90	558.0
159.00	473.0
159.90	879.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

161.00	1417.0
161.90	340.0
164.90	1160.0
165.90	781.0
167.00	6005.0
167.90	4214.0
169.00	554.0
170.00	220.0
170.90	259.0
171.90	492.0
173.00	615.0
173.90	1425.0
175.00	2172.0
176.00	698.0
177.00	1239.0
178.00	332.0
178.90	4194.0
179.90	2943.0
180.90	1018.0
181.90	170.0
183.80	348.0
185.00	2107.0
186.00	15730.0
187.00	4863.0
187.90	473.0
189.00	1049.0
189.90	191.0
191.00	537.0
191.90	1635.0
193.00	1701.0
194.00	317.0
194.90	213.0
196.00	2924.0
197.90	109872.0
198.90	8033.0
199.90	543.0
201.50	443.0
202.90	970.0
204.00	4275.0
205.00	7075.0
206.00	28168.0
207.00	5237.0
208.00	1216.0
208.90	566.0
209.90	422.0
210.90	1354.0
211.80	266.0
214.90	368.0
215.90	718.0
216.90	7784.0
218.00	1104.0
221.00	4025.0
221.90	1048.0
223.00	1758.0
224.00	15985.0
225.00	4234.0
225.90	400.0
226.90	7489.0
228.00	898.0
229.00	1353.0
229.90	220.0
230.90	560.0
233.00	151.0
234.00	487.0
234.90	487.0
235.90	465.0
236.90	468.0
238.90	354.0
239.90	230.0
240.90	256.0
241.90	877.0
242.90	778.0
244.00	12328.0
245.00	1645.0
245.90	2758.0
246.90	705.0
249.00	384.0
251.90	157.0
252.90	293.0
255.00	63360.0
256.00	9793.0
257.00	669.0
257.90	3597.0
258.90	701.0
264.90	1582.0
265.90	283.0
270.90	390.0
271.90	211.0
273.00	1752.0
274.00	4581.0
275.00	26424.0
276.00	3664.0
276.90	2436.0
277.90	453.0
281.10	234.0
283.00	247.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

283.90	160.0
285.00	496.0
293.00	602.0
294.00	234.0
296.00	7802.0
296.90	1046.0
303.00	883.0
303.90	258.0
314.00	307.0
314.90	963.0
316.00	476.0
320.90	216.0
323.00	2601.0
324.00	537.0
326.90	516.0
327.90	175.0
332.20	166.0
332.90	277.0
334.00	1649.0
335.00	273.0
340.90	265.0
345.90	463.0
351.90	739.0
353.00	511.0
354.00	574.0
365.00	3219.0
365.90	358.0
370.90	154.0
372.00	1026.0
373.00	243.0
382.90	278.0
402.00	562.0
403.00	574.0
404.00	231.0
420.90	477.0
422.00	443.0
423.00	2707.0
424.10	807.0
441.00	7041.0
442.00	45736.0
443.10	9967.0
444.00	1088.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

m/z	Abundance
36.90	224.0
38.10	559.0
39.10	3514.0
40.00	603.0
44.00	720.0
49.00	429.0
50.00	12328.0
51.10	45448.0
52.00	2259.0
55.00	198.0
56.00	1560.0
57.00	3433.0
58.00	174.0
61.00	717.0
62.00	670.0
63.00	2092.0
63.90	271.0
65.10	1102.0
69.00	60136.0
70.00	425.0
73.10	698.0
74.00	6605.0
75.00	10535.0
76.10	3475.0
77.10	69896.0
78.00	4572.0
79.00	5207.0
80.00	4189.0
81.00	5487.0
82.00	1209.0
83.00	1256.0
84.00	237.0
85.00	824.0
85.90	1552.0
86.90	724.0
87.90	351.0
91.00	1179.0
92.00	1150.0
93.00	9607.0
93.90	750.0
95.10	207.0
96.00	513.0
98.00	8204.0
99.00	5931.0
100.00	455.0
101.00	2944.0
102.00	181.0
102.90	1167.0
104.00	1973.0
105.00	1810.0
107.00	24744.0
108.00	3081.0
109.00	755.0
110.00	40280.0
111.00	6135.0
112.00	796.0
112.80	327.0
116.10	1394.0
117.00	23864.0
117.90	1787.0
119.00	198.0
120.00	212.0
122.00	1724.0
123.00	2711.0
123.90	1107.0
125.00	1067.0
127.00	78384.0
128.00	5754.0
129.00	32776.0
130.00	3048.0
131.00	614.0
132.00	288.0
133.00	218.0
134.00	1172.0
135.00	2965.0
135.90	1233.0
137.00	1594.0
138.00	325.0
139.90	423.0
141.00	4725.0
141.90	1535.0
142.80	913.0
144.10	338.0
145.00	206.0
146.00	812.0
147.00	2656.0
147.90	6253.0
148.90	1254.0
149.80	389.0
151.00	481.0
152.00	355.0
153.00	1363.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

154.00	1251.0
155.00	2498.0
155.90	3667.0
157.00	682.0
157.90	769.0
158.90	878.0
159.90	1509.0
160.90	2273.0
161.90	664.0
162.90	213.0
165.00	1858.0
166.00	1556.0
167.00	9911.0
168.00	5734.0
169.00	876.0
170.00	308.0
170.90	422.0
171.90	798.0
172.90	1220.0
174.00	2264.0
175.00	3988.0
175.90	1111.0
177.00	1979.0
178.00	648.0
178.90	7698.0
180.00	5299.0
181.00	2346.0
181.90	424.0
183.00	245.0
184.00	654.0
185.00	3716.0
186.00	26240.0
187.00	7956.0
187.90	869.0
188.90	1860.0
190.00	282.0
191.00	931.0
191.90	2303.0
193.00	3038.0
193.90	533.0
196.00	4907.0
197.90	184768.0
199.00	13530.0
200.00	1095.0
201.40	751.0
203.00	1299.0
204.00	7188.0
205.00	13332.0
206.00	51016.0
207.00	7579.0
208.00	2202.0
209.00	717.0
210.20	908.0
211.00	2018.0
213.00	180.0
214.90	546.0
216.00	1183.0
216.90	13797.0
218.00	1892.0
219.00	209.0
221.00	7142.0
221.90	2037.0
222.90	3238.0
224.00	28544.0
225.00	7476.0
225.90	903.0
227.00	13351.0
228.00	1892.0
228.90	2618.0
229.90	332.0
231.00	1022.0
231.90	189.0
232.90	273.0
233.90	833.0
235.00	852.0
236.00	565.0
236.90	1060.0
238.90	461.0
239.90	358.0
240.90	815.0
242.00	1461.0
243.00	1845.0
244.00	20224.0
245.00	2845.0
246.00	5034.0
246.90	1072.0
247.90	230.0
248.90	915.0
249.90	178.0
251.00	215.0
251.80	209.0
253.00	402.0
255.00	115024.0
256.00	16848.0
257.00	1388.0
257.90	6891.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

258.90	1075.0
259.90	208.0
260.90	188.0
264.00	250.0
264.90	2916.0
265.90	469.0
271.00	524.0
273.00	3466.0
274.00	9210.0
275.00	46976.0
276.00	6951.0
276.90	4985.0
277.90	839.0
281.00	284.0
282.00	209.0
282.90	562.0
284.10	340.0
285.00	888.0
286.10	170.0
289.00	162.0
290.00	169.0
290.90	218.0
291.90	206.0
293.00	1151.0
293.90	253.0
296.00	15717.0
297.00	2378.0
298.00	182.0
300.90	192.0
301.90	201.0
303.10	1701.0
304.00	586.0
307.90	189.0
309.90	163.0
312.90	173.0
314.00	739.0
314.90	1788.0
316.00	1137.0
321.00	452.0
321.90	266.0
323.00	4509.0
324.00	837.0
326.90	767.0
327.90	444.0
328.90	156.0
331.90	403.0
332.90	479.0
334.00	3361.0
334.90	706.0
341.00	498.0
341.90	199.0
345.90	1015.0
346.90	232.0
351.90	1538.0
353.00	876.0
354.00	1522.0
355.00	400.0
365.00	7741.0
365.90	1046.0
371.00	376.0
372.00	2448.0
373.10	482.0
382.90	701.0
383.90	152.0
390.00	229.0
391.00	219.0
391.90	161.0
401.90	756.0
403.00	1418.0
403.90	504.0
421.00	1034.0
422.00	1115.0
423.00	6941.0
423.90	1390.0
425.10	172.0
439.90	175.0
441.10	18528.0
442.10	105448.0
443.00	22160.0
444.10	1860.0
444.90	165.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

m/z	Abundance
38.10	481.0
39.10	2817.0
40.00	434.0
44.00	917.0
50.00	10788.0
51.00	36360.0
52.00	1881.0
55.00	172.0
56.00	1124.0
57.00	2801.0
60.90	589.0
62.00	820.0
63.00	1982.0
64.00	233.0
65.00	979.0
68.00	1053.0
69.00	51072.0
70.10	273.0
73.00	736.0
74.00	5478.0
75.00	8865.0
76.10	3123.0
77.10	58376.0
78.00	3938.0
79.00	4141.0
80.00	3081.0
81.00	4684.0
82.00	1029.0
83.00	933.0
83.90	162.0
85.00	704.0
86.00	1651.0
87.00	658.0
87.90	214.0
91.00	1281.0
92.00	1194.0
93.00	8160.0
94.00	402.0
95.90	277.0
96.90	189.0
98.00	6645.0
99.00	5107.0
100.10	416.0
101.00	2623.0
102.20	163.0
102.90	806.0
104.00	1748.0
105.00	1707.0
105.90	456.0
107.00	21216.0
108.00	3422.0
110.00	36712.0
111.00	5464.0
111.90	662.0
113.00	203.0
117.00	21120.0
118.00	1563.0
118.80	288.0
119.90	286.0
122.00	1556.0
123.00	2184.0
124.00	1121.0
125.00	935.0
127.00	70136.0
128.10	5442.0
129.00	29168.0
130.00	2235.0
130.90	381.0
132.00	257.0
132.90	339.0
133.90	940.0
135.00	2496.0
136.00	1049.0
137.00	1496.0
137.90	255.0
140.10	482.0
140.90	4223.0
141.90	1556.0
142.90	773.0
144.00	289.0
145.00	349.0
145.90	796.0
147.00	2393.0
148.00	5591.0
149.00	1060.0
149.90	278.0
151.00	485.0
151.90	378.0
152.90	1524.0
154.00	1052.0
155.00	2517.0
156.00	3772.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

157.10	650.0
157.90	706.0
158.90	543.0
160.00	1399.0
161.00	2050.0
161.90	583.0
164.00	210.0
164.90	1780.0
165.90	1362.0
167.00	9751.0
168.00	5690.0
168.90	870.0
170.10	282.0
171.00	505.0
171.90	959.0
173.00	1026.0
173.90	1755.0
175.00	3648.0
175.90	1064.0
176.90	1692.0
178.00	714.0
178.90	6641.0
180.00	4862.0
181.00	2277.0
182.00	336.0
182.90	159.0
184.00	545.0
185.00	3424.0
186.00	24952.0
187.00	7888.0
188.00	968.0
188.90	1794.0
189.90	437.0
190.90	1031.0
192.00	2347.0
193.00	2566.0
194.00	516.0
196.00	4917.0
198.00	183488.0
198.90	12043.0
200.00	1009.0
201.40	561.0
201.70	612.0
203.00	1392.0
204.00	7098.0
205.00	12037.0
206.00	49504.0
207.00	7311.0
208.00	2224.0
209.00	629.0
210.90	2236.0
214.90	651.0
216.00	1189.0
216.90	14487.0
218.00	1821.0
218.90	227.0
221.00	6478.0
221.90	1720.0
222.90	3079.0
224.00	28792.0
225.00	7101.0
227.00	13284.0
228.00	1763.0
228.90	2521.0
229.90	435.0
230.90	926.0
232.90	308.0
233.90	947.0
234.90	938.0
236.00	758.0
236.90	854.0
237.90	199.0
239.00	498.0
239.80	395.0
240.90	711.0
242.00	1807.0
243.10	1602.0
244.00	20960.0
245.00	2497.0
245.90	5479.0
246.90	856.0
248.00	198.0
248.90	697.0
251.00	205.0
251.90	290.0
252.90	513.0
255.00	117272.0
256.00	16648.0
257.00	1356.0
258.00	6801.0
259.00	1249.0
259.90	237.0
263.90	256.0
264.90	3132.0
265.90	441.0
271.00	595.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246

271.90	520.0
273.00	3199.0
274.00	10249.0
275.00	50176.0
276.00	6136.0
277.00	4903.0
278.00	937.0
279.10	209.0
281.00	274.0
282.00	184.0
283.00	606.0
283.90	331.0
285.00	809.0
285.90	166.0
288.90	247.0
289.90	186.0
291.00	171.0
292.00	349.0
293.00	1063.0
294.00	361.0
294.70	269.0
296.00	16800.0
297.00	2380.0
300.90	238.0
302.00	317.0
303.00	1639.0
303.90	569.0
308.00	238.0
308.90	161.0
310.00	237.0
313.10	201.0
314.00	804.0
314.90	1960.0
315.90	962.0
316.80	170.0
320.90	594.0
322.00	352.0
323.00	5140.0
324.00	909.0
327.00	1139.0
327.90	610.0
331.90	365.0
333.00	608.0
334.00	3550.0
335.00	978.0
335.90	156.0
340.90	810.0
342.00	251.0
346.00	1175.0
346.90	272.0
352.00	1551.0
352.90	1280.0
354.00	1735.0
354.90	378.0
364.90	7880.0
365.90	1236.0
371.00	466.0
372.00	2827.0
373.00	848.0
383.00	722.0
384.00	198.0
389.90	374.0
391.00	214.0
391.90	216.0
401.10	166.0
401.90	1062.0
402.90	1580.0
403.90	556.0
421.00	1207.0
422.10	1419.0
423.00	7488.0
424.00	1713.0
441.10	22728.0
442.00	136576.0
443.00	27480.0
444.10	2673.0

Instrument :
BNA_N
ClientSampleId :
DFTPP246