

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112719\
 Data File : BN008775.D
 Acq On : 28 Nov 2019 01:41
 Operator : JU
 Sample : PB124863BL
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK63

Quant Time: Nov 28 02:54:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN112719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 28 01:33:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	494778	20.00	ng/ul	0.00
18) Naphthalene-d8	10.35	136	2325437	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.21	164	1663404	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.96	188	3765277	20.00	ng/ul	0.00
77) Chrysene-d12	21.16	240	3607442	20.00	ng/ul	0.00
85) Perylene-d12	23.32	264	3937951	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.18	96	79980	7.07	ng/uL	0.00
5) Phenol-d5	6.78	99	1590956	31.34	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.93	67	1080478	34.51	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	1181114	33.36	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	1308189	32.54	ng/ul	0.00
19) Nitrobenzene-d5	8.72	128	639237	35.63	ng/ul	0.00
22) 2-Nitrophenol-d4	9.44	143	715653	36.93	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.97	165	1242273	32.19	ng/ul	0.00
29) 4-Chloroaniline-d4	10.48	131	1968759	40.14	ng/ul	0.00
43) Dimethylphthalate-d6	13.63	166	4888386	36.89	ng/ul	0.00
46) Acenaphthylene-d8	13.90	160	5233434	35.75	ng/ul	0.00
51) 4-Nitrophenol-d4	14.42	143	923471	34.36	ng/ul	0.00
57) Fluorene-d10	15.21	176	3947552	35.25	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.33	200	709432	32.61	ng/ul	0.00
70) Anthracene-d10	17.06	188	6074903	34.08	ng/ul	0.00
78) Pyrene-d10	19.36	212	7219200	35.87	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.19	264	7516253	36.58	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112719\
Data File : BN008775.D
Acq On : 28 Nov 2019 01:41
Operator : JU
Sample : PB124863BL
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK63

Quant Time: Nov 28 02:54:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN112719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 28 01:33:10 2019
Response via : Initial Calibration

