

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113021\
 Data File : BN017615.D
 Acq On : 30 Nov 2021 16:23
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Nov 30 17:29:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 30 16:26:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.773	152	24666	0.400	ng	0.00
7) Naphthalene-d8	10.547	136	98215	0.400	ng	# 0.00
13) Acenaphthene-d10	14.384	164	57321	0.400	ng	0.00
19) Phenanthrene-d10	17.128	188	113196	0.400	ng	0.00
29) Chrysene-d12	21.314	240	113747	0.400	ng	# 0.00
35) Perylene-d12	23.630	264	95884	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.366	112	95231	1.558	ng	0.00
5) Phenol-d6	6.943	99	103465	1.500	ng	0.00
8) Nitrobenzene-d5	8.912	82	84008	1.299	ng	0.00
11) 2-Methylnaphthalene-d10	12.142	152	289213	1.768	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	51946	2.950	ng	-0.01
15) 2-Fluorobiphenyl	13.014	172	370370	1.735	ng	0.00
27) Fluoranthene-d10	19.159	212	625118	1.795	ng	0.00
31) Terphenyl-d14	19.764	244	421864	1.688	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.328	88	30986	2.619	ng	99
3) n-Nitrosodimethylamine	3.642	42	39757	1.634	ng	97
6) bis(2-Chloroethyl)ether	7.201	93	115450	1.640	ng	97
9) Naphthalene	10.598	128	407983	1.637	ng	100
10) Hexachlorobutadiene	10.902	225	110940	2.258	ng	# 100
12) 2-Methylnaphthalene	12.213	142	274039	1.709	ng	99
16) Acenaphthylene	14.105	152	395373	1.729	ng	100
17) Acenaphthene	14.448	154	257651	1.631	ng	98
18) Fluorene	15.437	166	325756	1.711	ng	99
20) 4,6-Dinitro-2-methylph...	15.513	198	14836	2.814	ng	# 90
21) 4-Bromophenyl-phenylether	16.325	248	122440	1.970	ng	# 70
22) Hexachlorobenzene	16.447	284	139294	2.073	ng	# 93
23) Atrazine	16.593	200	77101	1.872	ng	98
24) Pentachlorophenol	16.788	266	55206	2.717	ng	100
25) Phenanthrene	17.165	178	516806	1.631	ng	100
26) Anthracene	17.262	178	478802	1.768	ng	100
28) Fluoranthene	19.188	202	614575	1.754	ng	98
30) Pyrene	19.551	202	612947	1.593	ng	100
32) Benzo(a)anthracene	21.292	228	612379	1.729	ng	99
33) Chrysene	21.346	228	613033	1.621	ng	98
34) Bis(2-ethylhexyl)phtha...	21.239	149	162844	1.193	ng	96
36) Indeno(1,2,3-cd)pyrene	26.005	276	533499	1.823	ng	96
37) Benzo(b)fluoranthene	22.928	252	551906	1.582	ng	98
38) Benzo(k)fluoranthene	22.976	252	551742	1.631	ng	# 98
39) Benzo(a)pyrene	23.527	252	488149	1.696	ng	# 97
40) Dibenzo(a,h)anthracene	26.022	278	439777	1.830	ng	97
41) Benzo(g,h,i)perylene	26.727	276	443646	1.771	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113021\
Data File : BN017615.D
Acq On : 30 Nov 2021 16:23
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Nov 30 17:29:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 30 16:26:30 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.773 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 24666
Ion Ratio Lower Upper
152 100
150 156.7 123.0 184.4
115 58.7 45.6 68.4



#2
1,4-Dioxane
Concen: 2.619 ng
RT: 3.328 min Scan# 38
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion: 88 Resp: 30986
Ion Ratio Lower Upper
88 100
43 31.1 24.7 37.1
58 78.1 63.0 94.6









#3
 n-Nitrosodimethylamine
 Concen: 1.634 ng
 RT: 3.642 min Scan# 71
 Delta R.T. -0.009 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion: 42 Resp: 39757




Ion	Ratio	Lower	Upper
42	100		
74	125.6	97.8	146.6
44	7.5	5.5	8.3




#4
 2-Fluorophenol
 Concen: 1.558 ng
 RT: 5.366 min Scan# 259
 Delta R.T. 0.000 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23

Tgt Ion: 112 Resp: 95231

Ion	Ratio	Lower	Upper
112	100		
64	64.1	47.6	71.4
63	33.1	24.6	37.0





#5

Phenol-d6

Concen: 1.500 ng

RT: 6.943 min Scan# 41

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion: 99 Resp: 103465

Ion Ratio Lower Upper

99 100

42 20.4 17.0 25.6

71 35.0 24.8 37.2



#6

bis(2-Chloroethyl)ether

Concen: 1.640 ng

RT: 7.201 min Scan# 458

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23




Tgt Ion: 93 Resp: 115450


Ion Ratio Lower Upper

93 100

63 78.6 65.8 98.6

95 32.1 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.547 min Scan# 755
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:136 Resp: 98215
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 8.9 4.0 6.0#
68 7.4 3.0 4.6#



#8
Nitrobenzene-d5
Concen: 1.299 ng
RT: 8.912 min Scan# 626
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion: 82 Resp: 84008
Ion Ratio Lower Upper
82 100
128 35.8 27.7 41.5
54 48.5 49.2 73.8#





#9

Naphthalene

Concen: 1.637 ng

RT: 10.598 min Scan# 759

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:128 Resp: 407983

Ion Ratio Lower Upper

128 100

129 10.8 8.8 13.2

127 13.2 10.6 15.8



#10

Hexachlorobutadiene

Concen: 2.258 ng

RT: 10.902 min Scan# 783

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23




Tgt Ion:225 Resp: 110940


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.1 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 1.768 ng

RT: 12.142 min Scan# 911

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23

Instrument :

BNA_N

ClientSampleId :





SSTDICC1.6

Tgt Ion:152 Resp: 289213

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.4



#12

2-Methylnaphthalene

Concen: 1.709 ng

RT: 12.213 min Scan# 989

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23




Tgt Ion:142 Resp: 274039


Ion Ratio Lower Upper

142 100

141 88.3 70.6 106.0

115 34.8 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.384 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 57321
Ion Ratio Lower Upper
164 100
162 101.3 81.4 122.2
160 46.2 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 2.950 ng
RT: 15.869 min Scan# 1320
Delta R.T. -0.013 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:330 Resp: 51946
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.5 28.9 43.3#









#15
2-Fluorobiphenyl
Concen: 1.735 ng
RT: 13.014 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 370370
Ion Ratio Lower Upper
172 100
171 34.4 28.7 43.1
170 23.0 19.4 29.2



#16
Acenaphthylene
Concen: 1.729 ng
RT: 14.105 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:152 Resp: 395373
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 1.631 ng
RT: 14.448 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 257651
Ion Ratio Lower Upper
154 100
153 110.6 91.2 136.8
152 54.7 44.3 66.5



#18
Fluorene
Concen: 1.711 ng
RT: 15.437 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:166 Resp: 325756
Ion Ratio Lower Upper
166 100
165 96.8 77.9 116.9
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.128 min Scan# 1423
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 113196
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.0 6.1 9.1



#20
4,6-Dinitro-2-methylphenol
Concen: 2.814 ng
RT: 15.513 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:198 Resp: 14836
Ion Ratio Lower Upper
198 100
182 10.3 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.970 ng
RT: 16.325 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 122440
Ion Ratio Lower Upper
248 100
250 97.8 71.8 107.8
141 40.0 70.9 106.3#



#22
Hexachlorobenzene
Concen: 2.073 ng
RT: 16.447 min Scan# 1367
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:284 Resp: 139294
Ion Ratio Lower Upper
284 100
142 27.6 28.2 42.2#
249 30.3 24.4 36.6









#23
 Atrazine
 Concen: 1.872 ng
 RT: 16.593 min Scan# 1379
 Delta R.T. 0.000 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 77101
 Ion Ratio Lower Upper
 200 100
 173 24.3 17.6 26.4
 215 49.6 40.2 60.4



#24
 Pentachlorophenol
 Concen: 2.717 ng
 RT: 16.788 min Scan# 1395
 Delta R.T. 0.000 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23

Tgt Ion:266 Resp: 55206
 Ion Ratio Lower Upper
 266 100
 264 62.3 49.8 74.8
 268 63.7 51.3 76.9









#25
Phenanthrene
Concen: 1.631 ng
RT: 17.165 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 516806
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.4 12.3 18.5



#26
Anthracene
Concen: 1.768 ng
RT: 17.262 min Scan# 1434
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:178 Resp: 478802
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.795 ng
RT: 19.159 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 625118
Ion Ratio Lower Upper
212 100
106 11.5 10.5 15.7
104 6.3 5.8 8.6



#28
Fluoranthene
Concen: 1.754 ng
RT: 19.188 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:202 Resp: 614575
Ion Ratio Lower Upper
202 100
101 9.9 8.9 13.3
203 17.5 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.314 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 113747
Ion Ratio Lower Upper
240 100
120 7.3 4.2 6.2#
236 26.2 20.0 30.0



#30
Pyrene
Concen: 1.593 ng
RT: 19.551 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:202 Resp: 612947
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 1.688 ng
 RT: 19.764 min Scan# 1848
 Delta R.T. 0.000 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 421864
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.7 8.5
 122 9.3 8.6 13.0



#32
 Benzo(a)anthracene
 Concen: 1.729 ng
 RT: 21.292 min Scan# 2018
 Delta R.T. -0.011 min
 Lab File: BN017615.D
 Acq: 30 Nov 2021 16:23

Tgt Ion:228 Resp: 612379
 Ion Ratio Lower Upper
 228 100
 226 27.6 21.3 31.9
 229 19.4 15.7 23.5









#33
Chrysene
Concen: 1.621 ng
RT: 21.346 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 613033
Ion Ratio Lower Upper
228 100
226 30.4 23.3 34.9
229 19.2 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.193 ng
RT: 21.239 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:149 Resp: 162844
Ion Ratio Lower Upper
149 100
167 29.3 21.8 32.6
279 4.4 3.4 5.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.630 min Scan# 24

Delta R.T. 0.000 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:264 Resp: 95884

Ion Ratio Lower Upper

264 100

260 24.6 19.4 29.2

265 23.5 18.6 28.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.823 ng

RT: 26.005 min Scan# 3174

Delta R.T. 0.004 min

Lab File: BN017615.D

Acq: 30 Nov 2021 16:23




Tgt Ion:276 Resp: 533499


Ion Ratio Lower Upper

276 100

138 20.8 19.7 29.5

277 25.0 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 1.582 ng
RT: 22.928 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 551906
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 8.0 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 1.631 ng
RT: 22.976 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:252 Resp: 551742
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 8.1 8.2 12.2#









#39
Benzo(a)pyrene
Concen: 1.696 ng
RT: 23.527 min Scan# 2450
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 488149
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.8 9.2 13.8#



#40
Dibenzo(a,h)anthracene
Concen: 1.830 ng
RT: 26.022 min Scan# 3179
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Tgt Ion:278 Resp: 439777
Ion Ratio Lower Upper
278 100
139 14.0 13.0 19.6
279 23.3 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 1.771 ng
RT: 26.727 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN017615.D
Acq: 30 Nov 2021 16:23

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 443646
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.4
138 17.9 17.1 25.7

