

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113021\
 Data File : BN017616.D
 Acq On : 30 Nov 2021 17:00
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Nov 30 17:30:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 30 16:26:30 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.773	152	24814	0.400	ng	0.00
7) Naphthalene-d8	10.547	136	96140	0.400	ng	# 0.00
13) Acenaphthene-d10	14.384	164	55319	0.400	ng	0.00
19) Phenanthrene-d10	17.128	188	111169	0.400	ng	0.00
29) Chrysene-d12	21.313	240	110614	0.400	ng	# 0.00
35) Perylene-d12	23.629	264	95438	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.366	112	184850	3.006	ng	0.00
5) Phenol-d6	6.943	99	203410	2.932	ng	0.00
8) Nitrobenzene-d5	8.912	82	189943	3.000	ng	0.00
11) 2-Methylnaphthalene-d10	12.142	152	541181	3.379	ng	0.00
14) 2,4,6-Tribromophenol	15.881	330	107080	6.301	ng	0.00
15) 2-Fluorobiphenyl	13.014	172	691959	3.359	ng	0.00
27) Fluoranthene-d10	19.158	212	1198950	3.506	ng	0.00
31) Terphenyl-d14	19.764	244	808235	3.326	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.328	88	58160	4.887	ng	Qvalue 100
3) n-Nitrosodimethylamine	3.642	42	77511	3.166	ng	94
6) bis(2-Chloroethyl)ether	7.201	93	219935	3.105	ng	95
9) Naphthalene	10.598	128	762204	3.124	ng	100
10) Hexachlorobutadiene	10.902	225	211305	4.393	ng	# 100
12) 2-Methylnaphthalene	12.213	142	505982	3.224	ng	98
16) Acenaphthylene	14.105	152	773639	3.505	ng	100
17) Acenaphthene	14.447	154	492240	3.229	ng	96
18) Fluorene	15.437	166	607935	3.309	ng	99
20) 4,6-Dinitro-2-methylph...	15.513	198	45359	8.761	ng	# 84
21) 4-Bromophenyl-phenylether	16.325	248	235938	3.865	ng	# 71
22) Hexachlorobenzene	16.447	284	260289	3.944	ng	# 93
23) Atrazine	16.593	200	165498	4.092	ng	97
24) Pentachlorophenol	16.787	266	124755	6.251	ng	100
25) Phenanthrene	17.165	178	975882	3.136	ng	100
26) Anthracene	17.262	178	917224	3.448	ng	100
28) Fluoranthene	19.188	202	1139727	3.313	ng	98
30) Pyrene	19.550	202	1177914	3.147	ng	100
32) Benzo(a)anthracene	21.292	228	1201148	3.488	ng	99
33) Chrysene	21.345	228	1160753	3.157	ng	98
34) Bis(2-ethylhexyl)phtha...	21.239	149	389028	2.930	ng	96
36) Indeno(1,2,3-cd)pyrene	25.998	276	1040713	3.572	ng	96
37) Benzo(b)fluoranthene	22.972	252	1082929	3.119	ng	98
38) Benzo(k)fluoranthene	22.972	252	1042482	3.095	ng	# 98
39) Benzo(a)pyrene	23.523	252	960134	3.351	ng	# 97
40) Dibenzo(a,h)anthracene	26.022	278	858172	3.588	ng	97
41) Benzo(g,h,i)perylene	26.724	276	878243	3.522	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113021\
Data File : BN017616.D
Acq On : 30 Nov 2021 17:00
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Nov 30 17:30:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN113021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 30 16:26:30 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.773 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:152 Resp: 24814
 Ion Ratio Lower Upper
 152 100
 150 155.9 123.0 184.4
 115 58.5 45.6 68.4



#2
 1,4-Dioxane
 Concen: 4.887 ng
 RT: 3.328 min Scan# 38
 Delta R.T. -0.000 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00

Tgt Ion: 88 Resp: 58160
 Ion Ratio Lower Upper
 88 100
 43 30.8 24.7 37.1
 58 78.3 63.0 94.6









#3
 n-Nitrosodimethylamine
 Concen: 3.166 ng
 RT: 3.642 min Scan# 71
 Delta R.T. -0.009 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion: 42 Resp: 77511
 Ion Ratio Lower Upper
 42 100
 74 128.8 97.8 146.6
 44 7.2 5.5 8.3



#4
 2-Fluorophenol
 Concen: 3.006 ng
 RT: 5.366 min Scan# 259
 Delta R.T. -0.000 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00

Tgt Ion: 112 Resp: 184850
 Ion Ratio Lower Upper
 112 100
 64 64.3 47.6 71.4
 63 33.0 24.6 37.0









#5
Phenol-d6
Concen: 2.932 ng
RT: 6.943 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 203410
Ion Ratio Lower Upper
99 100
42 20.4 17.0 25.6
71 35.1 24.8 37.2



#6
bis(2-Chloroethyl)ether
Concen: 3.105 ng
RT: 7.201 min Scan# 458
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion: 93 Resp: 219935
Ion Ratio Lower Upper
93 100
63 77.4 65.8 98.6
95 31.7 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.547 min Scan# 755
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	96140
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.8	13.2
54	9.4	4.0	6.0#
68	7.8	3.0	4.6#



#8
Nitrobenzene-d5
Concen: 3.000 ng
RT: 8.912 min Scan# 626
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:	82	Resp:	189943
Ion Ratio	Lower	Upper	
82	100		
128	35.0	27.7	41.5
54	49.3	49.2	73.8









#9
Naphthalene
Concen: 3.124 ng
RT: 10.598 min Scan# 759
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 762204
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.2 10.6 15.8



#10
Hexachlorobutadiene
Concen: 4.393 ng
RT: 10.902 min Scan# 783
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:225 Resp: 211305
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 3.379 ng

RT: 12.142 min Scan# 911

Delta R.T. -0.000 min

Lab File: BN017616.D

Acq: 30 Nov 2021 17:00

Instrument :

BNA_N

ClientSampleId :





SSTDICC3.2

Tgt Ion:152 Resp: 541181

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.4



#12

2-Methylnaphthalene

Concen: 3.224 ng

RT: 12.213 min Scan# 989

Delta R.T. -0.000 min

Lab File: BN017616.D

Acq: 30 Nov 2021 17:00




Tgt Ion:142 Resp: 505982


Ion Ratio Lower Upper

142 100

141 87.3 70.6 106.0

115 35.1 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.384 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 55319
Ion Ratio Lower Upper
164 100
162 101.9 81.4 122.2
160 46.5 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 6.301 ng
RT: 15.881 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:330 Resp: 107080
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.5 28.9 43.3#









#15
2-Fluorobiphenyl
Concen: 3.359 ng
RT: 13.014 min Scan# 1095
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 691959
Ion Ratio Lower Upper
172 100
171 34.4 28.7 43.1
170 23.1 19.4 29.2



#16
Acenaphthylene
Concen: 3.505 ng
RT: 14.105 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:152 Resp: 773639
Ion Ratio Lower Upper
152 100
151 19.3 15.5 23.3
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 3.229 ng
RT: 14.447 min Scan# 1208
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 492240
Ion Ratio Lower Upper
154 100
153 109.4 91.2 136.8
152 52.8 44.3 66.5



#18
Fluorene
Concen: 3.309 ng
RT: 15.437 min Scan# 1286
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:166 Resp: 607935
Ion Ratio Lower Upper
166 100
165 96.1 77.9 116.9
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.128 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 111169
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.3 6.1 9.1



#20
4,6-Dinitro-2-methylphenol
Concen: 8.761 ng
RT: 15.513 min Scan# 1292
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:198 Resp: 45359
Ion Ratio Lower Upper
198 100
182 7.7 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.865 ng
RT: 16.325 min Scan# 1357
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 235938
Ion Ratio Lower Upper
248 100
250 97.5 71.8 107.8
141 41.1 70.9 106.3#



#22
Hexachlorobenzene
Concen: 3.944 ng
RT: 16.447 min Scan# 1367
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:284 Resp: 260289
Ion Ratio Lower Upper
284 100
142 27.7 28.2 42.2#
249 30.3 24.4 36.6









#23
Atrazine
Concen: 4.092 ng
RT: 16.593 min Scan# 1379
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 165498
Ion Ratio Lower Upper
200 100
173 25.0 17.6 26.4
215 49.4 40.2 60.4



#24
Pentachlorophenol
Concen: 6.251 ng
RT: 16.787 min Scan# 1395
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:266 Resp: 124755
Ion Ratio Lower Upper
266 100
264 62.8 49.8 74.8
268 63.9 51.3 76.9









#25
Phenanthrene
Concen: 3.136 ng
RT: 17.165 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 975882
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.4 12.3 18.5



#26
Anthracene
Concen: 3.448 ng
RT: 17.262 min Scan# 1434
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:178 Resp: 917224
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.506 ng
RT: 19.158 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1198950
Ion Ratio Lower Upper
212 100
106 11.5 10.5 15.7
104 6.3 5.8 8.6



#28
Fluoranthene
Concen: 3.313 ng
RT: 19.188 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:202 Resp: 1139727
Ion Ratio Lower Upper
202 100
101 10.0 8.9 13.3
203 17.5 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 20
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 110614
Ion Ratio Lower Upper
240 100
120 7.2 4.2 6.2#
236 26.4 20.0 30.0



#30
Pyrene
Concen: 3.147 ng
RT: 19.550 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:202 Resp: 1177914
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.6 14.0 21.0









#31
 Terphenyl-d14
 Concen: 3.326 ng
 RT: 19.764 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 808235
 Ion Ratio Lower Upper
 244 100
 212 7.4 5.7 8.5
 122 9.4 8.6 13.0



#32
 Benzo(a)anthracene
 Concen: 3.488 ng
 RT: 21.292 min Scan# 2018
 Delta R.T. -0.011 min
 Lab File: BN017616.D
 Acq: 30 Nov 2021 17:00

Tgt Ion:228 Resp: 1201148
 Ion Ratio Lower Upper
 228 100
 226 27.5 21.3 31.9
 229 19.3 15.7 23.5









#33
Chrysene
Concen: 3.157 ng
RT: 21.345 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 1160753
Ion Ratio Lower Upper
228 100
226 30.3 23.3 34.9
229 19.2 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 2.930 ng
RT: 21.239 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:149 Resp: 389028
Ion Ratio Lower Upper
149 100
167 29.4 21.8 32.6
279 4.5 3.4 5.0









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.629 min Scan# 2480
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:264 Resp: 95438




Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.4	29.2
265	23.7	18.6	28.0




#36
Indeno(1,2,3-cd)pyrene
Concen: 3.572 ng
RT: 25.998 min Scan# 3172
Delta R.T. -0.004 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:276 Resp: 1040713

Ion	Ratio	Lower	Upper
276	100		
138	20.6	19.7	29.5
277	25.1	20.0	30.0









#37
Benzo(b)fluoranthene
Concen: 3.119 ng
RT: 22.972 min Scan# 2288
Delta R.T. 0.044 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1082929
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.0 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 3.095 ng
RT: 22.972 min Scan# 2288
Delta R.T. -0.004 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Tgt Ion:252 Resp: 1042482
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.0 8.2 12.2#









#39
Benzo(a)pyrene
Concen: 3.351 ng
RT: 23.523 min Scan# 24
Delta R.T. -0.004 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 960134
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.7 9.2 13.8#



#40
Dibenzo(a,h)anthracene
Concen: 3.588 ng
RT: 26.022 min Scan# 3179
Delta R.T. -0.000 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00


Tgt Ion:278 Resp: 858172
Ion Ratio Lower Upper
278 100
139 13.7 13.0 19.6
279 23.7 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 3.522 ng
RT: 26.724 min Scan# 31
Delta R.T. -0.004 min
Lab File: BN017616.D
Acq: 30 Nov 2021 17:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:276 Resp: 878243
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.4
138 17.6 17.1 25.7

