

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113023\
 Data File : BN028921.D
 Acq On : 30 Nov 2023 22:12
 Operator : MA/JU
 Sample : PB156727BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156727BL

Quant Time: Dec 01 00:29:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 00:28:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.811	152	4781	0.400	ng	0.00
7) Naphthalene-d8	10.594	136	13550	0.400	ng	0.00
13) Acenaphthene-d10	14.428	164	5455	0.400	ng	0.00
19) Phenanthrene-d10	17.184	188	13809	0.400	ng	# 0.00
29) Chrysene-d12	21.374	240	3971	0.400	ng	0.00
35) Perylene-d12	23.714	264	174	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	5257	0.363	ng	0.00
5) Phenol-d6	7.046	99	4613	0.287	ng	0.00
8) Nitrobenzene-d5	8.993	82	3593	0.267	ng	0.01
11) 2-Methylnaphthalene-d10	12.182	152	7145	0.329	ng	0.00
14) 2,4,6-Tribromophenol	15.930	330	1277	0.346	ng	0.00
15) 2-Fluorobiphenyl	13.060	172	9769	0.415	ng	-0.01
27) Fluoranthene-d10	19.213	212	9906	0.273	ng	0.00
31) Terphenyl-d14	19.822	244	8365	0.615	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113023\
Data File : BN028921.D
Acq On : 30 Nov 2023 22:12
Operator : MA/JU
Sample : PB156727BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156727BL

Quant Time: Dec 01 00:29:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 00:28:34 2023
Response via : Initial Calibration

