

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113023\
 Data File : BN028923.D
 Acq On : 30 Nov 2023 23:23
 Operator : MA/JU
 Sample : 05350-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE117D1-20231107

Quant Time: Dec 01 00:30:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 00:28:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	3659	0.400	ng	0.00
7) Naphthalene-d8	10.594	136	9219	0.400	ng	0.00
13) Acenaphthene-d10	14.428	164	5114	0.400	ng	0.00
19) Phenanthrene-d10	17.171	188	10651	0.400	ng	#-0.01
29) Chrysene-d12	21.374	240	6435	0.400	ng	# 0.00
35) Perylene-d12	23.705	264	6258	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1239	0.112	ng	0.00
5) Phenol-d6	7.045	99	641	0.052	ng	0.00
8) Nitrobenzene-d5	8.982	82	2055	0.224	ng	0.00
11) 2-Methylnaphthalene-d10	12.181	152	3935	0.266	ng	0.00
14) 2,4,6-Tribromophenol	15.930	330	873	0.253	ng	0.00
15) 2-Fluorobiphenyl	13.060	172	5406	0.245	ng	-0.01
27) Fluoranthene-d10	19.209	212	10002	0.357	ng	0.00
31) Terphenyl-d14	19.817	244	9673	0.481	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.369	88	1272	0.355	ng	# 78
28) Fluoranthene	19.241	202	1359	0.037	ng	95
30) Pyrene	19.603	202	1357	0.122	ng	# 95
32) Benzo(a)anthracene	21.356	228	1265	0.046	ng	# 89
33) Chrysene	21.409	228	1214	0.045	ng	# 85
36) Indeno(1,2,3-cd)pyrene	26.102	276	1202	0.044	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN113023\
Data File : BN028923.D
Acq On : 30 Nov 2023 23:23
Operator : MA/JU
Sample : 05350-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE117D1-20231107

Quant Time: Dec 01 00:30:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 00:28:34 2023
Response via : Initial Calibration

