

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120118\
 Data File : BN003660.D
 Acq On : 01 Dec 2018 14:03
 Operator : JU/SJ
 Sample : J6096-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4116

Manual Integrations
 APPROVED

mohammad
 12/3/2018 4:59:45 PM

Quant Time: Dec 03 01:44:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 03 01:15:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5435	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	22861	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	12913	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	27379m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	23831m	0.40	ng/ul	0.00
20) Perylene-d12	23.70	264	19191m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	12890	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	28678m	0.36	ng/ul	0.00
Target Compounds						
8) Acenaphthene	14.55	153	6124	0.120	ng/ul#	Ovalue 23

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120118\
Data File : BN003660.D
Acq On : 01 Dec 2018 14:03
Operator : JU/SJ
Sample : J6096-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4116

Quant Time: Dec 03 01:44:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 03 01:15:58 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
12/3/2018 4:59:45 PM

