


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120118\
 Data File : BN003652.D
 Acq On : 01 Dec 2018 09:21
 Operator : JU/SJ
 Sample : DFTPP58
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP58

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 03 03:30:35 2018



AutoFind: Scans 2489, 2490, 2491; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.9	53002	PASS
68	69	0.00	2	1.8	1106	PASS
69	198	0.00	100	29.3	60085	PASS
70	69	0.00	2	0.4	239	PASS
127	198	10	80	42.1	86264	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	204992	PASS
199	198	5	9	6.7	13754	PASS
275	198	10	60	31.4	64344	PASS
365	198	1	100	4.1	8393	PASS
441	198	0.01	100	14.0	28650	PASS
442	198	50	100	90.1	184725	PASS
443	442	15	24	19.0	35066	PASS

Scan 2482 (17.586 min): BN003652.D
DFTPP58

m/z	Abundance
44.00	2402.0
207.00	327.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	840.0
39.00	5027.0
39.90	425.0
44.00	1982.0
49.10	617.0
50.00	17184.0
51.00	67592.0
52.00	3451.0
55.00	367.0
56.00	2109.0
57.00	4670.0
61.00	930.0
62.00	1023.0
63.00	2935.0
64.00	418.0
65.00	1414.0
68.10	1414.0
69.00	75864.0
70.00	353.0
73.00	682.0
74.00	7625.0
75.00	12251.0
76.00	3964.0
77.00	85880.0
78.00	5992.0
79.00	5666.0
80.00	4499.0
81.00	6374.0
82.00	1439.0
83.00	1462.0
85.00	1106.0
86.00	1792.0
87.00	853.0
90.90	1384.0
92.00	1691.0
93.00	9753.0
93.90	677.0
96.00	515.0
98.00	7785.0
99.00	6683.0
100.00	651.0
101.00	3623.0
103.00	1147.0
104.00	2277.0
105.00	2089.0
106.00	666.0
107.00	27520.0
108.00	4467.0
109.00	879.0
110.00	54384.0
111.00	8027.0
112.00	960.0
116.00	1714.0
117.00	21064.0
118.00	1743.0
119.90	407.0
122.00	1943.0
123.00	2986.0
124.00	1245.0
124.90	1283.0
127.00	105272.0
128.00	7998.0
129.00	40304.0
130.00	3393.0
131.00	681.0
132.00	350.0
134.00	1216.0
135.00	3310.0
136.00	1431.0
137.00	1755.0
137.80	416.0
139.90	516.0
141.00	5408.0
142.00	1653.0
143.00	1223.0
143.90	339.0
146.00	1013.0
147.00	2669.0
148.00	5739.0
149.00	1425.0
149.90	374.0
150.90	613.0
153.00	1519.0
154.00	1248.0
155.00	2840.0
156.00	4236.0
157.00	928.0
158.00	1002.0
158.90	752.0
160.00	1663.0
161.00	2433.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

162.00	705.0
165.00	2044.0
166.00	1602.0
167.00	10524.0
168.00	6419.0
169.00	1007.0
170.10	388.0
170.90	480.0
171.90	979.0
173.00	1313.0
174.00	2248.0
175.00	4239.0
176.00	1302.0
176.90	1972.0
178.00	646.0
179.00	8089.0
180.00	5838.0
181.00	2673.0
182.00	422.0
184.00	676.0
185.00	3920.0
186.00	30352.0
187.00	9133.0
188.00	962.0
189.00	1990.0
189.90	310.0
190.90	871.0
192.00	2403.0
193.00	2851.0
194.00	626.0
195.00	396.0
196.00	6850.0
198.00	233280.0
199.00	16226.0
200.00	1237.0
201.50	1118.0
203.00	1439.0
204.00	7486.0
205.00	13837.0
206.00	59064.0
207.10	8272.0
208.00	2006.0
209.00	690.0
210.40	953.0
211.00	2310.0
215.00	621.0
216.00	1272.0
217.00	15159.0
218.00	1869.0
221.00	11182.0
223.00	3462.0
224.00	33448.0
225.00	8665.0
226.00	961.0
227.00	13006.0
228.00	1862.0
229.00	2792.0
230.00	407.0
231.00	1124.0
234.00	878.0
235.00	974.0
236.00	529.0
237.00	1280.0
239.00	472.0
240.00	493.0
241.00	738.0
242.00	1717.0
243.00	1887.0
244.10	27056.0
245.10	3541.0
246.00	5134.0
247.00	987.0
248.90	939.0
253.00	544.0
255.00	139968.0
256.00	20856.0
257.00	1816.0
258.00	7292.0
259.00	1227.0
265.00	3247.0
266.00	470.0
271.00	304.0
271.90	323.0
273.00	4269.0
274.00	11638.0
275.00	68912.0
276.00	9210.0
277.00	5385.0
278.00	1015.0
283.00	566.0
284.00	483.0
285.00	918.0
293.00	1230.0
294.00	345.0
296.00	18144.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

297.00	2601.0
301.00	305.0
302.00	308.0
303.10	2012.0
304.00	632.0
314.00	960.0
315.00	1994.0
316.00	1299.0
321.00	606.0
322.10	380.0
323.10	6109.0
324.00	1100.0
326.90	1077.0
328.00	714.0
331.90	444.0
333.00	710.0
334.00	4203.0
335.00	1086.0
341.00	763.0
346.00	1302.0
352.00	1909.0
353.00	1420.0
354.10	1930.0
355.10	411.0
365.00	8493.0
366.00	1143.0
371.10	520.0
372.10	3313.0
373.10	894.0
383.00	968.0
390.00	488.0
391.00	307.0
402.00	1352.0
403.00	1898.0
404.00	741.0
421.10	1437.0
422.00	1385.0
423.10	10908.0
424.10	3185.0
425.00	346.0
441.10	26328.0
442.10	169344.0
443.10	33208.0
444.10	3240.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	711.0
39.00	4433.0
40.00	381.0
44.00	1995.0
49.10	485.0
50.00	15516.0
51.00	60800.0
52.00	3234.0
55.00	301.0
56.00	1907.0
57.00	4597.0
61.00	858.0
62.00	989.0
63.00	2695.0
64.00	385.0
65.00	1226.0
68.10	1272.0
69.00	69472.0
70.00	365.0
73.00	544.0
74.00	7042.0
75.00	11084.0
76.00	3886.0
77.00	78928.0
78.00	5580.0
79.00	5515.0
80.00	4167.0
81.00	5780.0
82.00	1470.0
82.90	1418.0
85.00	962.0
86.00	1709.0
87.00	777.0
87.90	320.0
91.00	1407.0
92.00	1525.0
93.00	9325.0
94.00	630.0
96.00	445.0
98.00	7110.0
99.00	6116.0
100.00	579.0
101.00	3461.0
103.00	1326.0
104.00	2202.0
105.00	1877.0
106.00	698.0
107.00	25808.0
108.00	4210.0
109.10	875.0
110.00	50864.0
111.00	7810.0
112.00	886.0
116.00	1488.0
117.00	20536.0
118.00	1584.0
120.00	308.0
122.00	1812.0
123.00	2828.0
124.00	1309.0
125.00	1260.0
127.00	100792.0
128.00	7739.0
129.00	39008.0
130.00	3357.0
131.00	556.0
131.90	316.0
134.00	1085.0
135.00	3253.0
136.00	1250.0
137.00	1781.0
137.90	400.0
140.00	526.0
141.00	5063.0
142.00	1601.0
143.00	1115.0
144.00	338.0
146.00	913.0
147.00	2611.0
148.00	5515.0
148.90	1286.0
149.90	338.0
151.10	684.0
153.00	1709.0
154.00	1252.0
155.00	2830.0
156.00	4628.0
157.00	1016.0
158.00	998.0
159.00	710.0
160.00	1566.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

161.00	2413.0
162.00	734.0
164.00	311.0
165.00	1932.0
166.00	1609.0
167.00	10121.0
168.00	5690.0
169.00	1037.0
169.90	318.0
170.90	484.0
172.00	970.0
173.00	1229.0
174.00	2261.0
175.00	4302.0
176.00	1267.0
177.00	1732.0
177.90	635.0
179.00	8020.0
180.00	5586.0
181.00	2614.0
182.00	425.0
184.00	635.0
185.00	3851.0
186.00	31864.0
187.00	8772.0
188.00	1029.0
188.90	1956.0
190.00	327.0
191.10	891.0
192.00	2576.0
193.00	2643.0
194.00	566.0
195.00	404.0
196.00	7293.0
198.00	245440.0
199.00	16031.0
200.00	1211.0
201.40	1207.0
203.00	1526.0
204.00	8278.0
205.00	14181.0
206.10	61920.0
207.10	8283.0
208.00	2007.0
209.00	713.0
210.30	1029.0
211.00	2597.0
215.00	677.0
216.00	1323.0
217.00	16055.0
218.00	1999.0
221.10	11016.0
223.00	3613.0
224.00	34376.0
225.00	9047.0
226.10	965.0
227.00	13583.0
228.00	1948.0
229.00	2956.0
230.00	400.0
231.00	1296.0
234.00	931.0
235.00	1012.0
236.00	630.0
237.00	1276.0
238.90	551.0
239.90	442.0
241.00	821.0
242.00	1924.0
243.00	2110.0
244.10	28728.0
245.10	3789.0
246.00	5547.0
247.00	1072.0
249.00	1166.0
252.90	595.0
255.00	155840.0
256.00	22384.0
257.00	1720.0
258.00	8289.0
259.00	1213.0
265.00	3355.0
265.90	459.0
271.00	321.0
272.00	398.0
273.00	4750.0
274.00	12564.0
275.00	77232.0
276.00	9781.0
277.00	5657.0
278.00	1098.0
283.10	619.0
284.00	513.0
285.00	1001.0
293.00	1336.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

294.10	335.0
296.00	20008.0
297.00	2726.0
301.90	365.0
303.00	2430.0
304.00	712.0
307.90	335.0
314.00	978.0
315.00	2169.0
316.00	1520.0
321.00	682.0
322.00	333.0
323.00	7124.0
324.00	1323.0
327.00	1355.0
328.00	646.0
331.90	475.0
333.00	706.0
334.00	4819.0
335.00	1277.0
341.00	879.0
346.00	1393.0
352.10	2194.0
353.00	1668.0
354.00	2317.0
355.00	508.0
365.00	9913.0
366.00	1447.0
371.00	643.0
372.00	4246.0
373.00	929.0
383.00	1165.0
383.90	372.0
390.10	555.0
391.00	416.0
402.00	1734.0
403.00	2395.0
404.00	872.0
421.10	1924.0
422.00	1847.0
423.10	13826.0
424.10	3456.0
425.00	444.0
441.10	34440.0
442.10	221056.0
443.10	41912.0
444.10	4056.0

Instrument :

BNA_N

ClientSampleId :

DFTPP58

m/z	Abundance
38.00	380.0
39.00	2181.0
39.90	334.0
44.00	2426.0
50.00	7939.0
51.00	30616.0
52.00	1463.0
55.90	926.0
57.00	2165.0
61.00	400.0
62.00	463.0
63.00	1364.0
65.00	643.0
68.10	633.0
69.00	34920.0
73.00	357.0
74.00	3668.0
75.00	5659.0
76.00	1924.0
77.00	39672.0
78.00	2814.0
79.00	2725.0
80.00	1941.0
81.00	2895.0
82.00	757.0
83.00	694.0
85.00	491.0
86.00	934.0
86.90	424.0
90.90	710.0
92.00	805.0
93.00	4860.0
94.00	343.0
98.00	3760.0
99.00	3245.0
101.00	1777.0
102.90	640.0
103.90	1107.0
105.00	1094.0
106.00	359.0
107.00	13747.0
108.00	2109.0
109.20	513.0
110.00	26136.0
111.00	3915.0
111.90	469.0
116.00	727.0
117.00	10602.0
117.90	716.0
121.90	902.0
123.00	1473.0
124.00	679.0
125.00	618.0
127.00	52728.0
128.00	4097.0
129.00	20400.0
130.00	1585.0
131.00	332.0
133.90	535.0
135.00	1790.0
135.90	675.0
136.90	892.0
140.00	309.0
141.00	2607.0
142.00	883.0
143.00	534.0
146.00	553.0
147.00	1461.0
148.00	3040.0
149.00	775.0
151.00	400.0
152.90	899.0
154.00	722.0
155.00	1608.0
156.00	2345.0
157.00	508.0
157.90	515.0
159.00	420.0
160.00	895.0
161.00	1334.0
161.90	385.0
164.90	1083.0
166.00	811.0
167.00	5633.0
168.00	3158.0
169.00	581.0
171.90	522.0
173.00	682.0
174.00	1292.0
175.00	2191.0
176.00	642.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

177.00	1207.0
177.90	337.0
179.00	4433.0
180.00	3105.0
181.00	1608.0
183.90	378.0
185.00	2207.0
186.00	17704.0
187.00	4871.0
188.00	463.0
188.90	1053.0
191.00	530.0
192.00	1478.0
193.00	1667.0
194.00	327.0
196.00	3796.0
198.00	136256.0
199.00	9007.0
200.00	632.0
201.50	620.0
203.00	956.0
204.00	4575.0
205.00	8051.0
206.00	35600.0
207.00	4635.0
208.00	1181.0
208.90	387.0
210.20	520.0
211.00	1460.0
215.00	367.0
216.00	643.0
217.00	9075.0
218.00	1082.0
221.00	6428.0
223.00	2104.0
224.00	19992.0
225.10	5182.0
226.00	635.0
227.00	7899.0
228.00	1039.0
229.00	1846.0
231.00	681.0
234.00	553.0
235.00	617.0
236.00	362.0
237.00	840.0
239.00	398.0
241.00	487.0
242.00	1182.0
243.10	1156.0
244.10	17464.0
245.00	2304.0
246.00	3369.0
247.00	636.0
248.90	674.0
253.00	368.0
255.00	93128.0
256.00	13001.0
257.00	992.0
258.00	5200.0
259.00	745.0
265.00	2072.0
273.00	3005.0
274.00	7905.0
275.00	46888.0
276.00	6148.0
277.00	3665.0
278.00	631.0
283.00	407.0
283.90	303.0
285.10	647.0
293.00	834.0
295.10	357.0
296.00	12707.0
297.00	1747.0
303.00	1628.0
304.00	475.0
314.00	615.0
315.00	1613.0
316.00	901.0
321.00	501.0
323.10	4832.0
324.00	898.0
326.90	941.0
328.00	430.0
332.00	317.0
333.00	506.0
334.00	3237.0
335.00	934.0
341.00	595.0
346.00	1134.0
352.00	1562.0
353.00	1105.0
354.00	1718.0
365.00	6773.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

366.00	847.0
371.00	402.0
372.10	2824.0
373.00	600.0
383.00	804.0
390.00	374.0
402.00	1139.0
403.10	1631.0
404.10	786.0
421.00	1351.0
422.00	1250.0
423.10	9959.0
424.00	2274.0
425.10	306.0
441.10	25184.0
442.10	163776.0
443.10	30080.0
444.10	2714.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58