

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028959.D
 Acq On : 01 Dec 2023 22:24
 Operator : MA/JU
 Sample : 05357-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW1-8-20231108

Quant Time: Dec 01 23:17:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	2420	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	7139	0.400	ng	0.00
13) Acenaphthene-d10	14.418	164	4349	0.400	ng	0.00
19) Phenanthrene-d10	17.171	188	9534	0.400	ng	0.00
29) Chrysene-d12	21.365	240	7804	0.400	ng	# 0.00
35) Perylene-d12	23.699	264	7187	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.413	112	583	0.080	ng	0.00
5) Phenol-d6	7.009	99	409	0.050	ng	0.00
8) Nitrobenzene-d5	8.971	82	1770	0.249	ng	0.00
11) 2-Methylnaphthalene-d10	12.170	152	2732	0.238	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	229	0.078	ng	0.00
15) 2-Fluorobiphenyl	13.049	172	8814	0.469	ng	0.00
27) Fluoranthene-d10	19.204	212	8271	0.330	ng	0.00
31) Terphenyl-d14	19.813	244	7138	0.350	ng	0.00
Target Compounds						Qvalue
16) Acenaphthylene	14.140	152	698	0.030	ng	# 95
30) Pyrene	19.599	202	215	0.098	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
Data File : BN028959.D
Acq On : 01 Dec 2023 22:24
Operator : MA/JU
Sample : 05357-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW1-8-20231108

Quant Time: Dec 01 23:17:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:11:55 2023
Response via : Initial Calibration

