

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028964.D
 Acq On : 02 Dec 2023 01:26
 Operator : MA/JU
 Sample : 05358-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D6-20231108

Quant Time: Dec 02 03:16:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	2355	0.400	ng	0.00
7) Naphthalene-d8	10.584	136	7035	0.400	ng	# 0.00
13) Acenaphthene-d10	14.418	164	4195	0.400	ng	0.00
19) Phenanthrene-d10	17.171	188	9251	0.400	ng	# 0.00
29) Chrysene-d12	21.365	240	7902	0.400	ng	# 0.00
35) Perylene-d12	23.693	264	7295	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.413	112	781	0.109	ng	0.00
5) Phenol-d6	7.010	99	470	0.059	ng	0.00
8) Nitrobenzene-d5	8.971	82	1807	0.258	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	2905	0.257	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	689	0.243	ng	0.00
15) 2-Fluorobiphenyl	13.049	172	8481	0.468	ng	0.00
27) Fluoranthene-d10	19.204	212	8243	0.339	ng	0.00
31) Terphenyl-d14	19.808	244	7525	0.358	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.333	88	10153	4.402	ng	Qvalue 96
6) bis(2-Chloroethyl)ether	7.255	93	239	0.036	ng	97
16) Acenaphthylene	14.140	152	620	0.028	ng	# 92
30) Pyrene	19.599	202	113	0.096	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
Data File : BN028964.D
Acq On : 02 Dec 2023 01:26
Operator : MA/JU
Sample : 05358-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D6-20231108

Quant Time: Dec 02 03:16:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:11:55 2023
Response via : Initial Calibration

