

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028991.D
 Acq On : 02 Dec 2023 19:59
 Operator : MA/JU
 Sample : 05404-11
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT101D2-20231109

Quant Time: Dec 03 08:15:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	2279	0.400	ng	-0.01
7) Naphthalene-d8	10.573	136	6519	0.400	ng	#-0.01
13) Acenaphthene-d10	14.407	164	4208	0.400	ng	-0.01
19) Phenanthrene-d10	17.159	188	9054	0.400	ng	-0.01
29) Chrysene-d12	21.356	240	7529	0.400	ng	#-0.02
35) Perylene-d12	23.678	264	6869	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	731	0.106	ng	0.00
5) Phenol-d6	6.995	99	457	0.060	ng	-0.01
8) Nitrobenzene-d5	8.950	82	2219	0.342	ng	-0.02
11) 2-Methylnaphthalene-d10	12.159	152	3191	0.305	ng	-0.01
14) 2,4,6-Tribromophenol	15.905	330	722	0.254	ng	-0.01
15) 2-Fluorobiphenyl	13.038	172	5706	0.314	ng	-0.01
27) Fluoranthene-d10	19.190	212	9443	0.397	ng	-0.02
31) Terphenyl-d14	19.799	244	8929	0.410	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.333	88	6518	2.920	ng	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
Data File : BN028991.D
Acq On : 02 Dec 2023 19:59
Operator : MA/JU
Sample : 05404-11
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT101D2-20231109

Quant Time: Dec 03 08:15:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:11:55 2023
Response via : Initial Calibration

