

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028992.D
 Acq On : 02 Dec 2023 20:35
 Operator : MA/JU
 Sample : 05404-12
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D1-20231109

Quant Time: Dec 03 08:15:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	1965	0.400	ng	-0.01
7) Naphthalene-d8	10.573	136	5651	0.400	ng	#-0.01
13) Acenaphthene-d10	14.407	164	3731	0.400	ng	-0.01
19) Phenanthrene-d10	17.159	188	8199	0.400	ng	-0.01
29) Chrysene-d12	21.356	240	7362	0.400	ng	#-0.02
35) Perylene-d12	23.678	264	6973	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	507	0.085	ng	-0.01
5) Phenol-d6	6.995	99	348	0.053	ng	-0.01
8) Nitrobenzene-d5	8.950	82	1806	0.321	ng	-0.02
11) 2-Methylnaphthalene-d10	12.159	152	2721	0.300	ng	-0.01
14) 2,4,6-Tribromophenol	15.905	330	630	0.250	ng	-0.01
15) 2-Fluorobiphenyl	13.038	172	4837	0.300	ng	-0.01
27) Fluoranthene-d10	19.190	212	8521	0.395	ng	-0.02
31) Terphenyl-d14	19.799	244	7987	0.388	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.333	88	1118	0.581	ng	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
Data File : BN028992.D
Acq On : 02 Dec 2023 20:35
Operator : MA/JU
Sample : 05404-12
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D1-20231109

Quant Time: Dec 03 08:15:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:11:55 2023
Response via : Initial Calibration

