

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028968.D
 Acq On : 02 Dec 2023 03:50
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 02 04:18:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	53	0.00
2	1,4-Dioxane	0.400	0.403	-0.8	54	0.00
3	n-Nitrosodimethylamine	0.400	0.375	6.3	60	0.00
4 S	2-Fluorophenol	0.400	0.444	-11.0	59	0.00
5 S	Phenol-d6	0.400	0.494	-23.5	68	0.01
6	bis(2-Chloroethyl)ether	0.400	0.430	-7.5	62	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	58	0.01
8 S	Nitrobenzene-d5	0.400	0.436	-9.0	66	0.01
9	Naphthalene	0.400	0.405	-1.3	59	0.01
10	Hexachlorobutadiene	0.400	0.376	6.0	54	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.434	-8.5	64	0.01
12	2-Methylnaphthalene	0.400	0.428	-7.0	63	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	61	0.01
14 S	2,4,6-Tribromophenol	0.400	0.381	4.8	61	0.01
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	62	0.01
16	Acenaphthylene	0.400	0.419	-4.7	67	0.01
17	Acenaphthene	0.400	0.399	0.3	63	0.01
18	Fluorene	0.400	0.450	-12.5	74	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.280	30.0#	50	0.01
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	62	0.01
22	Hexachlorobenzene	0.400	0.384	4.0	64	0.02
23	Atrazine	0.400	0.389	2.8	63	0.02
24	Pentachlorophenol	0.400	0.344	14.0	59	0.01
25	Phenanthrene	0.400	0.402	-0.5	70	0.01
26	Anthracene	0.400	0.399	0.3	68	0.01
27 SURR	Fluoranthene-d10	0.400	0.443	-10.7	80	0.00
28	Fluoranthene	0.400	0.450	-12.5	80	0.01
29 I	Chrysene-d12	0.400	0.400	0.0	84	0.02
30	Pyrene	0.400	0.345	13.8	83	0.01
31 S	Terphenyl-d14	0.400	0.409	-2.2	95	0.01
32	Benzo(a)anthracene	0.400	0.389	2.8	85	0.02
33	Chrysene	0.400	0.383	4.3	82	0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.120	70.0#	47	0.02
35 I	Perylene-d12	0.400	0.400	0.0	63	0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.408	-2.0	65	0.03
37	Benzo(b)fluoranthene	0.400	0.403	-0.8	66	0.02
38	Benzo(k)fluoranthene	0.400	0.401	-0.3	63	0.02
39 C	Benzo(a)pyrene	0.400	0.380	5.0	62	0.02
40	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	67	0.03
41	Benzo(g,h,i)perylene	0.400	0.403	-0.8	63	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4