

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120123\
 Data File : BN028987.D
 Acq On : 02 Dec 2023 17:00
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 03 08:14:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 23:11:55 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	53	0.00
2	1,4-Dioxane	0.392	0.369	5.9	51	0.00
3	n-Nitrosodimethylamine	0.513	0.539	-5.1	75	0.00
4 S	2-Fluorophenol	1.212	1.308	-7.9	58	0.00
5 S	Phenol-d6	1.343	1.584	-17.9	66	0.00
6	bis(2-Chloroethyl)ether	1.130	1.161	-2.7	59	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	57	0.01
8 S	Nitrobenzene-d5	0.398	0.438	-10.1	66	0.01
9	Naphthalene	1.271	1.287	-1.3	58	0.00
10	Hexachlorobutadiene	0.260	0.257	1.2	56	0.01
11 SURR	2-Methylnaphthalene-d10	0.642	0.694	-8.1	63	0.00
12	2-Methylnaphthalene	0.799	0.867	-8.5	63	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
14 S	2,4,6-Tribromophenol	0.270	0.261	3.3	61	0.00
15 S	2-Fluorobiphenyl	1.728	1.728	0.0	61	0.00
16	Acenaphthylene	2.119	2.138	-0.9	64	0.00
17	Acenaphthene	1.406	1.415	-0.6	63	0.00
18	Fluorene	1.780	2.004	-12.6	73	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	65	0.00
20	4,6-Dinitro-2-methylphenol	0.088	0.068	22.7	54	0.01
21	4-Bromophenyl-phenylether	0.275	0.261	5.1	62	0.01
22	Hexachlorobenzene	0.338	0.322	4.7	63	0.01
23	Atrazine	0.241	0.231	4.1	61	0.01
24	Pentachlorophenol	0.157	0.134	14.6	58	0.00
25	Phenanthrene	1.321	1.296	1.9	67	0.00
26	Anthracene	1.221	1.180	3.4	65	0.01
27 SURR	Fluoranthene-d10	1.052	1.176	-11.8	79	0.00
28	Fluoranthene	1.373	1.543	-12.4	79	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
30	Pyrene	2.391	1.918	19.8	81	0.00
31 S	Terphenyl-d14	1.413	1.183	16.3	94	0.00
32	Benzo(a)anthracene	1.701	1.601	5.9	81	0.00
33	Chrysene	1.670	1.626	2.6	81	0.00
34	Bis(2-ethylhexyl)phthalate	2.450	1.262	48.5#	40#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	61	0.00
36	Indeno(1,2,3-cd)pyrene	1.743	1.795	-3.0	63	0.00
37	Benzo(b)fluoranthene	1.636	1.726	-5.5	66	0.00
38	Benzo(k)fluoranthene	1.544	1.571	-1.7	62	0.00
39 C	Benzo(a)pyrene	1.462	1.442	1.4	62	0.00
40	Dibenzo(a,h)anthracene	1.349	1.414	-4.8	65	0.01
41	Benzo(g,h,i)perylene	1.517	1.547	-2.0	62	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4