

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120125\
 Data File : BN038239.D
 Acq On : 01 Dec 2025 23:17
 Operator : RC/JU
 Sample : Q3690-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW8-SP100-20251118

Quant Time: Dec 02 01:11:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 28 21:04:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.768	152	9532	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	28839	0.400	ng	#-0.01
13) Acenaphthene-d10	14.420	164	13442	0.400	ng	-0.01
19) Phenanthrene-d10	17.173	188	20421	0.400	ng	#-0.01
29) Chrysene-d12	21.366	240	15899	0.400	ng	0.00
35) Perylene-d12	23.671	264	16933	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	3375	0.151	ng	0.00
5) Phenol-d6	6.930	99	2323	0.083	ng	0.00
8) Nitrobenzene-d5	8.918	82	7856	0.341	ng	0.00
11) 2-Methylnaphthalene-d10	12.162	152	12538	0.306	ng	0.00
14) 2,4,6-Tribromophenol	15.920	330	1666	0.394	ng	-0.01
15) 2-Fluorobiphenyl	13.041	172	20218	0.389	ng	-0.01
27) Fluoranthene-d10	19.211	212	19349	0.437	ng	0.00
31) Terphenyl-d14	19.810	244	15401	0.419	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	551	0.055	ng	# 20
34) Bis(2-ethylhexyl)phtha...	21.286	149	5994	0.203	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120125\
Data File : BN038239.D
Acq On : 01 Dec 2025 23:17
Operator : RC/JU
Sample : Q3690-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251118

Quant Time: Dec 02 01:11:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 28 21:04:10 2025
Response via : Initial Calibration

