

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120225\
 Data File : BN038265.D
 Acq On : 02 Dec 2025 21:17
 Operator : RC/JU
 Sample : Q3688-17
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 M-29

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/03/2025
 Supervised By :Sohil Jodhani 12/03/2025

Quant Time: Dec 03 01:29:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN112625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 01 16:36:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.762	152	7498	0.400 ng/ul	0.00
4) Naphthalene-d8	10.562	136	24549	0.400 ng/ul	-0.01
9) Acenaphthene-d10	14.428	164	12936	0.400 ng/ul	0.00
13) Phenanthrene-d10	17.179	188	18613m	0.400 ng/ul	0.00
17) Chrysene-d12	21.368	240	11286m	0.400 ng/ul	0.00
23) Perylene-d12	23.674	264	10119m	0.400 ng/ul	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.214	96	48488	6.163 ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.162	152	12958	0.370 ng/ul	-0.01
18) Fluoranthene-d10	19.211	212	17813m	0.406 ng/ul	0.00
Target Compounds					
					Qvalue
2) 1,4-Dioxane	3.252	88	290	0.034 ng/ul#	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120225\
Data File : BN038265.D
Acq On : 02 Dec 2025 21:17
Operator : RC/JU
Sample : Q3688-17
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
M-29

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/03/2025
Supervised By :Sohil Jodhani 12/03/2025

Quant Time: Dec 03 01:29:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN112625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 01 16:36:13 2025
Response via : Initial Calibration

