

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120318\
 Data File : BN003680.D
 Acq On : 03 Dec 2018 15:23
 Operator : JU/SJ
 Sample : J6096-14
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4122

Quant Time: Dec 03 16:23:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 03 03:30:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	4650	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	18062	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	9608	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	22815	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	21954	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	19559	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	9431	0.34	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	22085	0.33	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.68	128	41454	0.817	ng/ul	100
7) Acenaphthylene	14.19	152	33207	0.768	ng/ul	100
8) Acenaphthene	14.54	153	27393	0.723	ng/ul	100
9) Fluorene	15.52	166	37791	0.855	ng/ul	99
11) Pentachlorophenol	16.86	266	10219	2.487	ng/ul	99
12) Phenanthrene	17.26	178	52219	0.697	ng/ul	99
13) Anthracene	17.35	178	27132	0.441	ng/ul	99
17) Pyrene	19.64	202	123341	1.566	ng/ul	99
18) Benzo(a)anthracene	21.38	228	49238	0.710	ng/ul	99
19) Chrysene	21.43	228	76621	0.966	ng/ul	100
22) Benzo(k)fluoranthene	23.06	252	34760	0.441	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	26.06	276	50106	0.624	ng/ul#	91
26) Benzo(g,h,i)perylene	26.79	276	56895	0.815	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120318\
Data File : BN003680.D
Acq On : 03 Dec 2018 15:23
Operator : JU/SJ
Sample : J6096-14
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4122

Quant Time: Dec 03 16:23:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 03 03:30:35 2018
Response via : Initial Calibration

