

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
 Data File : BN029012.D
 Acq On : 03 Dec 2023 09:46
 Operator : MA/JU
 Sample : 05404-04
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT205S1-20231109

Quant Time: Dec 04 00:42:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 00:28:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.782	152	2020	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	5897	0.400	ng	0.00
13) Acenaphthene-d10	14.407	164	3640	0.400	ng	0.00
19) Phenanthrene-d10	17.146	188	7769	0.400	ng	#-0.01
29) Chrysene-d12	21.347	240	6336	0.400	ng	# 0.00
35) Perylene-d12	23.672	264	6082	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.399	112	641	0.105	ng	0.00
5) Phenol-d6	6.988	99	452	0.067	ng	0.00
8) Nitrobenzene-d5	8.950	82	1876	0.320	ng	0.00
11) 2-Methylnaphthalene-d10	12.151	152	2675	0.283	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	588	0.239	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	4826	0.307	ng	0.00
27) Fluoranthene-d10	19.190	212	7950	0.389	ng	0.00
31) Terphenyl-d14	19.794	244	7394	0.406	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.340	88	528	0.267	ng	# 53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
Data File : BN029012.D
Acq On : 03 Dec 2023 09:46
Operator : MA/JU
Sample : 05404-04
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT205S1-20231109

Quant Time: Dec 04 00:42:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 00:28:48 2023
Response via : Initial Calibration

