

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
 Data File : BN029022.D
 Acq On : 03 Dec 2023 15:44
 Operator : MA/JU
 Sample : 05404-30
 Misc :
 ALS Vial : 76 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE122D1-20231110

Quant Time: Dec 04 00:44:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 00:28:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	1970	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	5422	0.400	ng	# 0.00
13) Acenaphthene-d10	14.396	164	3450	0.400	ng	-0.01
19) Phenanthrene-d10	17.146	188	7189	0.400	ng	-0.01
29) Chrysene-d12	21.347	240	5882	0.400	ng	# 0.00
35) Perylene-d12	23.672	264	5593	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	534	0.089	ng	0.00
5) Phenol-d6	6.988	99	342	0.052	ng	0.00
8) Nitrobenzene-d5	8.950	82	1327	0.246	ng	0.00
11) 2-Methylnaphthalene-d10	12.151	152	2002	0.230	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	512	0.220	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	3937	0.264	ng	0.00
27) Fluoranthene-d10	19.185	212	5899	0.312	ng	0.00
31) Terphenyl-d14	19.794	244	5607	0.358	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.326	88	4003	2.075	ng	# 88
30) Pyrene	19.580	202	412	0.103	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
Data File : BN029022.D
Acq On : 03 Dec 2023 15:44
Operator : MA/JU
Sample : 05404-30
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE122D1-20231110

Quant Time: Dec 04 00:44:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 00:28:48 2023
Response via : Initial Calibration

