

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
 Data File : BN029035.D
 Acq On : 04 Dec 2023 00:39
 Operator : MA/JU
 Sample : 05428-01
 Misc :
 ALS Vial : 90 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D2-20231113

Quant Time: Dec 04 02:23:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 00:28:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	1836	0.400	ng	0.00
7) Naphthalene-d8	10.551	136	5102	0.400	ng	#-0.01
13) Acenaphthene-d10	14.396	164	3059	0.400	ng	-0.01
19) Phenanthrene-d10	17.146	188	6483	0.400	ng	-0.01
29) Chrysene-d12	21.347	240	5270	0.400	ng	# 0.00
35) Perylene-d12	23.664	264	5081	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	578	0.104	ng	0.00
5) Phenol-d6	6.980	99	347	0.056	ng	0.00
8) Nitrobenzene-d5	8.950	82	1425	0.281	ng	0.00
11) 2-Methylnaphthalene-d10	12.147	152	2083	0.254	ng	0.00
14) 2,4,6-Tribromophenol	15.893	330	410	0.198	ng	-0.01
15) 2-Fluorobiphenyl	13.028	172	3673	0.278	ng	-0.01
27) Fluoranthene-d10	19.181	212	6067	0.356	ng	0.00
31) Terphenyl-d14	19.794	244	5726	0.388	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.326	88	5453	3.033	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
Data File : BN029035.D
Acq On : 04 Dec 2023 00:39
Operator : MA/JU
Sample : 05428-01
Misc :
ALS Vial : 90 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D2-20231113

Quant Time: Dec 04 02:23:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 00:28:48 2023
Response via : Initial Calibration

