

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
 Data File : BN029036.D
 Acq On : 04 Dec 2023 02:26
 Operator : MA/JU
 Sample : 05428-02
 Misc :
 ALS Vial : 91 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D1-20231113

Quant Time: Dec 04 03:00:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 00:28:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.768	152	1540	0.400	ng	-0.01
7) Naphthalene-d8	10.552	136	4141	0.400	ng	#-0.01
13) Acenaphthene-d10	14.396	164	2561	0.400	ng	-0.01
19) Phenanthrene-d10	17.147	188	5293	0.400	ng	-0.01
29) Chrysene-d12	21.347	240	4287	0.400	ng	# 0.00
35) Perylene-d12	23.667	264	4142	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	499	0.107	ng	-0.01
5) Phenol-d6	6.981	99	300	0.058	ng	0.00
8) Nitrobenzene-d5	8.939	82	1425	0.346	ng	-0.01
11) 2-Methylnaphthalene-d10	12.144	152	2000	0.301	ng	0.00
14) 2,4,6-Tribromophenol	15.893	330	411	0.238	ng	-0.01
15) 2-Fluorobiphenyl	13.028	172	3673	0.332	ng	-0.01
27) Fluoranthene-d10	19.181	212	5630	0.404	ng	0.00
31) Terphenyl-d14	19.789	244	5332	0.423	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.319	88	9703	6.434	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120323\
Data File : BN029036.D
Acq On : 04 Dec 2023 02:26
Operator : MA/JU
Sample : 05428-02
Misc :
ALS Vial : 91 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D1-20231113

Quant Time: Dec 04 03:00:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN113023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 00:28:48 2023
Response via : Initial Calibration

