

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN120518\
 Data File : BN003714.D
 Acq On : 05 Dec 2018 16:45
 Operator : JU/SJ
 Sample : J6170-01DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9M00DL

Quant Time: Dec 05 17:47:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 05 12:03:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3927	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	15795	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	8515	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	18631	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	15778	0.40	ng/ul	0.00
20) Perylene-d12	23.70	264	16216	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.23	152	1601	0.07	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	3700	0.07	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.68	128	29289	0.660	ng/ul	Qvalue 100
5) 2-Methylnaphthalene	12.30	142	1771	0.057	ng/ul	99
8) Acenaphthene	14.53	153	1650	0.049	ng/ul	98
9) Fluorene	15.52	166	1750	0.045	ng/ul#	97
12) Phenanthrene	17.25	178	2749	0.045	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN120518\
Data File : BN003714.D
Acq On : 05 Dec 2018 16:45
Operator : JU/SJ
Sample : J6170-01DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9M00DL

Quant Time: Dec 05 17:47:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 05 12:03:03 2018
Response via : Initial Calibration

