

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120518\
 Data File : BN003725.D
 Acq On : 05 Dec 2018 23:40
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.487

Quant Time: Dec 06 02:57:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 06 02:51:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
3	Naphthalene	1.124	1.161	-3.3	128	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.596	3.1	120	0.00
5	2-Methylnaphthalene	0.785	0.788	-0.4	125	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
7	Acenaphthylene	1.801	1.846	-2.5	105	0.00
8 C	Acenaphthene	1.578	1.625	-3.0	104	0.00
9	Fluorene	1.840	1.796	2.4	98	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
11	Pentachlorophenol	0.072	0.077	-6.9	109	0.00
12	Phenanthrene	1.313	1.355	-3.2	90	0.00
13	Anthracene	1.079	1.110	-2.9	91	0.00
14 SURR	Fluoranthene-d10	1.170	1.074	8.2	81	0.00
15 C	Fluoranthene	1.550	1.474	4.9	84	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
17	Pyrene	1.435	1.733	-20.8#	84	0.00
18	Benzo(a)anthracene	1.263	1.328	-5.1	82	0.00
19	Chrysene	1.446	1.465	-1.3	77	0.00
20 I	Perylene-d12	1.000	1.000	0.0	70	0.00
21	Benzo(b)fluoranthene	1.655	1.685	-1.8	71	0.00
22	Benzo(k)fluoranthene	1.612	1.669	-3.5	73	0.00
23 C	Benzo(a)pyrene	1.433	1.505	-5.0	74	0.00
24	Indeno(1,2,3-cd)pyrene	1.643	1.674	-1.9	71	0.00
25	Dibenzo(a,h)anthracene	1.322	1.331	-0.7	71	0.00
26	Benzo(g,h,i)perylene	1.428	1.416	0.8	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0