

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN120518\
 Data File : BN003728.D
 Acq On : 06 Dec 2018 11:36
 Operator : JU/SJ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.488

Quant Time: Dec 06 12:17:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 06 02:51:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	160#	0.00
3	Naphthalene	1.124	1.170	-4.1	172#	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.590	4.1	159#	0.00
5	2-Methylnaphthalene	0.785	0.790	-0.6	168#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
7	Acenaphthylene	1.801	1.842	-2.3	141	0.00
8 C	Acenaphthene	1.578	1.627	-3.1	140	0.00
9	Fluorene	1.840	1.806	1.8	133	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
11	Pentachlorophenol	0.072	0.082	-13.9	158#	0.00
12	Phenanthrene	1.313	1.360	-3.6	122	0.00
13	Anthracene	1.079	1.119	-3.7	125	0.00
14 SURR	Fluoranthene-d10	1.170	1.105	5.6	114	0.00
15 C	Fluoranthene	1.550	1.527	1.5	119	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
17	Pyrene	1.435	1.689	-17.7	119	0.00
18	Benzo(a)anthracene	1.263	1.363	-7.9	123	0.00
19	Chrysene	1.446	1.453	-0.5	112	0.00
20 I	Perylene-d12	1.000	1.000	0.0	104	0.00
21	Benzo(b)fluoranthene	1.655	1.673	-1.1	106	0.00
22	Benzo(k)fluoranthene	1.612	1.651	-2.4	108	0.00
23 C	Benzo(a)pyrene	1.433	1.501	-4.7	110	0.01
24	Indeno(1,2,3-cd)pyrene	1.643	1.635	0.5	104	0.00
25	Dibenzo(a,h)anthracene	1.322	1.290	2.4	103	0.00
26	Benzo(g,h,i)perylene	1.428	1.380	3.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0