

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120519\
 Data File : BN008852.D
 Acq On : 05 Dec 2019 17:50
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 06 03:16:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 04 15:42:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.384	0.376	2.1	100	0.00
3	n-Nitrosodimethylamine	0.531	0.467	12.1	100	0.00
4 S	2-Fluorophenol	1.083	1.042	3.8	98	0.00
5 S	Phenol-d6	1.475	1.451	1.6	103	0.00
6	bis(2-Chloroethyl)ether	1.327	1.296	2.3	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.353	0.345	2.3	105	0.00
9	Naphthalene	1.067	1.077	-0.9	105	-0.01
10	Hexachlorobutadiene	0.210	0.212	-1.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.664	0.671	-1.1	106	0.00
12	2-Methylnaphthalene	0.786	0.793	-0.9	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.218	0.202	7.3	101	0.00
15 S	2-Fluorobiphenyl	1.499	1.510	-0.7	105	0.00
16	Acenaphthylene	1.866	1.783	4.4	104	0.00
17	Acenaphthene	1.184	1.173	0.9	105	0.00
18	Fluorene	1.570	1.579	-0.6	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.234	0.232	0.9	107	0.00
21	Hexachlorobenzene	0.270	0.271	-0.4	105	0.00
22	Pentachlorophenol	0.108	0.090	16.7	102	0.00
23	Phenanthrene	1.221	1.214	0.6	105	0.00
24	Anthracene	1.121	1.082	3.5	105	0.00
25 SURR	Fluoranthene-d10	1.211	1.194	1.4	105	0.00
26	Fluoranthene	1.470	1.463	0.5	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.396	1.389	0.5	104	0.00
29 S	Terphenyl-d14	0.921	0.915	0.7	104	0.00
30	Benzo(a)anthracene	1.446	1.431	1.0	104	0.00
31	Chrysene	1.403	1.408	-0.4	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.964	0.776	19.5	93	0.00
33	Indeno(1,2,3-cd)pyrene	1.733	1.804	-4.1	108	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.329	1.294	2.6	103	-0.01
36	Benzo(k)fluoranthene	1.305	1.341	-2.8	110	0.00
37 C	Benzo(a)pyrene	1.253	1.242	0.9	106	-0.02
38	Dibenzo(a,h)anthracene	1.280	1.302	-1.7	107	-0.01
39	Benzo(a,h,i)perylene	1.296	1.313	-1.3	107	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120519\
Data File : BN008852.D
Acq On : 05 Dec 2019 17:50
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 06 03:16:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 04 15:42:09 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			