

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120618\
 Data File : BN003735.D
 Acq On : 06 Dec 2018 17:39
 Operator : JU/SJ
 Sample : J6077-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9Y38

Manual Integrations
 APPROVED

mohammad
 12/10/2018 12:20:14 PM

Quant Time: Dec 07 03:29:17 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 07 03:02:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5138	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	21811	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	12770	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	27536m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	23054m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	17295m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	11472	0.34	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	25341m	0.31	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.68	128	145458	2.374	ng/ul	99
5) 2-Methylnaphthalene	12.29	142	136813	3.198	ng/ul	100
8) Acenaphthene	14.53	153	3395	0.067	ng/ul	97
9) Fluorene	15.52	166	4125	0.070	ng/ul#	97
12) Phenanthrene	17.25	178	26242	0.290	ng/ul	99
17) Pyrene	19.63	202	3102m	0.038	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120618\
Data File : BN003735.D
Acq On : 06 Dec 2018 17:39
Operator : JU/SJ
Sample : J6077-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9Y38

Manual Integrations
APPROVED

mohammad
12/10/2018 12:20:14 PM

Quant Time: Dec 07 03:29:17 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 07 03:02:08 2018
Response via : Initial Calibration

