


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120618\
 Data File : BN003729.D
 Acq On : 06 Dec 2018 13:57
 Operator : JU/SJ
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP62

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 07 04:08:50 2018



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.3	54160	PASS
68	69	0.00	2	1.9	1194	PASS
69	198	0.00	100	31.2	61936	PASS
70	69	0.00	2	0.4	242	PASS
127	198	10	80	43.1	85509	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	198528	PASS
199	198	5	9	6.7	13255	PASS
275	198	10	60	30.8	61226	PASS
365	198	1	100	3.9	7840	PASS
441	198	0.01	100	13.2	26248	PASS
442	198	50	100	85.0	168824	PASS
443	442	15	24	19.5	32960	PASS

m/z	Abundance
38.00	623.0
39.00	4199.0
39.90	362.0
44.00	739.0
49.10	417.0
50.00	14177.0
51.00	54472.0
52.00	2910.0
55.00	415.0
56.00	1664.0
57.00	3791.0
61.00	747.0
62.00	759.0
63.00	2415.0
64.00	387.0
65.00	1069.0
68.00	1164.0
69.00	61880.0
70.00	337.0
73.00	533.0
74.00	6006.0
75.00	9903.0
76.10	3335.0
77.00	69152.0
78.00	4795.0
79.00	4689.0
80.00	3463.0
81.00	4912.0
82.00	1180.0
83.00	1139.0
85.00	859.0
86.00	1604.0
87.00	720.0
91.00	1181.0
92.00	1364.0
93.00	7571.0
94.00	521.0
95.90	395.0
98.00	6025.0
99.00	5333.0
100.00	496.0
101.00	2820.0
103.00	1030.0
104.00	1739.0
105.00	1651.0
106.10	594.0
107.00	20944.0
108.00	3236.0
109.10	684.0
110.00	41472.0
111.00	6497.0
112.00	678.0
116.10	1091.0
117.00	16087.0
118.00	1350.0
120.00	311.0
122.00	1525.0
123.00	2240.0
124.00	992.0
125.00	966.0
127.00	81304.0
128.00	6282.0
129.00	31056.0
130.00	2570.0
131.00	559.0
134.00	973.0
135.00	2529.0
136.00	1045.0
137.00	1384.0
140.00	385.0
141.00	4054.0
142.00	1431.0
143.00	824.0
146.00	643.0
147.00	2042.0
148.00	4331.0
149.00	858.0
151.10	540.0
151.80	403.0
152.90	1270.0
154.00	911.0
155.00	2107.0
156.00	3447.0
157.00	679.0
157.90	795.0
159.00	554.0
160.00	1341.0
161.00	1781.0
162.00	494.0
165.00	1558.0
166.00	1280.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

167.00	7482.0
168.00	4350.0
169.00	725.0
170.90	379.0
171.90	648.0
173.00	900.0
174.00	1671.0
175.00	2988.0
176.00	968.0
177.00	1464.0
177.90	557.0
179.00	5614.0
180.00	4399.0
181.00	1903.0
182.00	308.0
184.00	498.0
185.00	2876.0
186.00	22528.0
187.00	6645.0
188.00	780.0
189.00	1328.0
191.00	669.0
191.90	1831.0
193.00	2252.0
194.00	507.0
195.10	302.0
196.00	5417.0
198.00	172608.0
199.00	11904.0
200.00	866.0
201.50	955.0
202.90	1010.0
204.10	5545.0
205.00	9959.0
206.00	42160.0
207.10	5918.0
208.00	1568.0
208.90	545.0
211.00	1636.0
215.00	505.0
216.00	868.0
217.00	10666.0
218.00	1435.0
221.10	7552.0
223.00	2600.0
224.00	23016.0
225.00	6049.0
226.10	788.0
227.00	9399.0
228.00	1346.0
229.00	2036.0
230.00	317.0
231.00	864.0
234.00	614.0
235.00	659.0
236.10	345.0
237.00	953.0
239.00	399.0
240.00	313.0
240.90	551.0
242.00	1325.0
243.10	1445.0
244.00	19304.0
245.00	2618.0
246.00	3542.0
247.00	728.0
249.00	725.0
253.00	379.0
255.00	99360.0
256.00	14948.0
257.00	1254.0
258.00	5428.0
259.00	912.0
265.00	2239.0
265.90	310.0
272.00	311.0
273.00	3126.0
274.00	8119.0
275.00	49352.0
276.00	7080.0
277.00	3730.0
278.00	688.0
283.00	457.0
283.90	305.0
285.00	721.0
293.00	905.0
296.00	12365.0
297.00	1832.0
303.00	1531.0
304.00	372.0
314.00	686.0
315.00	1490.0
316.10	822.0
321.00	413.0
323.10	4216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

324.00	858.0
327.00	764.0
327.90	423.0
332.10	350.0
333.00	390.0
334.00	2751.0
335.00	825.0
341.00	454.0
346.00	1015.0
352.00	1383.0
353.00	982.0
354.00	1359.0
355.00	350.0
365.00	5670.0
366.00	831.0
372.00	2324.0
373.10	599.0
382.90	600.0
402.00	979.0
403.00	1236.0
404.00	530.0
421.10	934.0
422.10	1126.0
423.10	7159.0
424.10	2299.0
425.00	312.0
441.10	17208.0
442.10	112744.0
443.10	21440.0
444.10	2150.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

m/z	Abundance
38.00	799.0
39.00	4760.0
40.00	384.0
44.00	765.0
49.10	457.0
50.00	15969.0
51.00	63336.0
52.00	3151.0
55.00	388.0
56.00	1986.0
57.00	4604.0
61.00	913.0
62.00	1053.0
63.00	2750.0
64.00	382.0
65.00	1459.0
68.10	1412.0
69.00	72736.0
70.00	390.0
73.00	584.0
74.00	7304.0
75.00	11179.0
76.00	3829.0
77.00	82160.0
78.00	5705.0
79.00	5784.0
80.00	4260.0
81.00	5818.0
82.00	1520.0
83.00	1332.0
85.00	1095.0
86.00	1685.0
86.90	822.0
87.90	324.0
91.00	1411.0
92.00	1576.0
93.00	9261.0
94.00	595.0
96.00	415.0
98.00	7357.0
99.00	6054.0
100.00	571.0
101.00	3474.0
103.00	1098.0
104.00	2172.0
105.00	1907.0
106.00	722.0
107.00	26040.0
108.00	3988.0
109.10	797.0
110.00	51888.0
111.00	7698.0
112.00	883.0
116.00	1502.0
117.00	19824.0
118.00	1388.0
120.00	342.0
122.00	1887.0
123.00	2759.0
124.00	1281.0
125.00	1179.0
127.00	100800.0
128.00	7712.0
129.00	38248.0
130.00	3210.0
131.00	560.0
132.00	366.0
134.00	1158.0
135.00	3179.0
136.00	1250.0
137.00	1536.0
138.00	341.0
140.00	539.0
141.00	4976.0
142.00	1603.0
143.00	1160.0
144.00	302.0
146.00	937.0
147.00	2542.0
148.00	5543.0
149.00	1182.0
150.00	321.0
151.20	678.0
151.90	432.0
153.00	1583.0
154.00	1240.0
155.00	2929.0
156.00	4264.0
157.10	859.0
158.00	966.0
158.90	726.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

160.00	1662.0
161.00	2481.0
161.90	741.0
165.00	1921.0
166.00	1630.0
167.00	9782.0
168.00	5640.0
169.00	814.0
169.90	353.0
170.90	485.0
172.00	964.0
173.00	1156.0
174.00	2117.0
175.00	4043.0
176.00	1343.0
177.00	2067.0
178.00	626.0
179.00	7982.0
180.00	5606.0
181.00	2461.0
182.00	407.0
184.00	580.0
185.00	3851.0
186.00	30440.0
187.00	8880.0
188.00	987.0
189.00	1790.0
191.00	833.0
192.00	2512.0
193.00	2586.0
194.00	588.0
195.10	339.0
196.00	7275.0
198.00	235584.0
199.00	15804.0
200.00	1189.0
201.50	1308.0
203.00	1529.0
204.00	7814.0
205.00	13776.0
206.00	60184.0
207.00	8348.0
208.00	1885.0
209.00	631.0
210.20	993.0
211.00	2333.0
215.00	665.0
216.00	1286.0
217.00	15447.0
218.00	1940.0
221.00	11343.0
223.00	3586.0
224.00	33688.0
225.00	8763.0
226.10	987.0
227.00	12978.0
228.00	2035.0
229.00	2836.0
230.00	492.0
231.00	1275.0
234.00	838.0
235.00	1003.0
236.00	627.0
237.00	1312.0
239.00	600.0
240.00	507.0
241.00	764.0
242.00	1970.0
243.10	2068.0
244.10	28776.0
245.10	3757.0
246.00	5111.0
247.00	1023.0
249.00	1119.0
252.90	548.0
255.00	146816.0
256.00	21552.0
257.00	1803.0
258.00	7876.0
259.00	1264.0
265.00	3198.0
265.90	468.0
271.00	339.0
271.90	415.0
273.00	4550.0
274.00	12055.0
275.00	72952.0
276.00	9568.0
277.00	5678.0
278.00	1032.0
283.00	611.0
283.90	452.0
285.00	1045.0
288.90	309.0
292.00	332.0

Instrument :

BNA_N

ClientSampleId :

DFTPP62

293.00	1296.0
294.10	364.0
296.00	19000.0
297.00	2844.0
302.00	442.0
303.00	2259.0
304.00	609.0
314.00	954.0
315.00	2073.0
316.00	1442.0
321.00	567.0
322.10	403.0
323.10	6741.0
324.00	1236.0
327.00	1183.0
327.90	640.0
332.10	497.0
332.90	668.0
334.00	4417.0
335.00	1210.0
341.00	805.0
345.90	1353.0
347.00	331.0
352.00	1987.0
353.10	1450.0
354.00	2092.0
355.00	390.0
365.00	9140.0
366.00	1225.0
371.00	511.0
372.10	3812.0
373.10	969.0
383.00	941.0
390.00	508.0
391.00	370.0
402.00	1433.0
403.00	2275.0
404.00	898.0
421.00	1574.0
422.00	1723.0
423.10	12799.0
424.10	3291.0
425.00	405.0
441.10	30464.0
442.10	196032.0
443.10	39400.0
444.10	3553.0

Instrument :

BNA_N

ClientSampleId :

DFTPP62

m/z	Abundance
38.00	519.0
39.00	3387.0
40.00	327.0
44.00	734.0
49.00	325.0
50.00	11354.0
51.00	44672.0
52.00	2236.0
56.00	1324.0
57.00	3338.0
61.00	616.0
61.90	622.0
63.00	2005.0
65.00	930.0
68.10	1007.0
69.00	51192.0
73.00	514.0
74.00	5193.0
75.00	8124.0
76.00	2776.0
77.00	58792.0
78.00	3929.0
79.00	3852.0
80.00	3125.0
81.00	3931.0
82.00	994.0
83.00	1112.0
85.00	716.0
86.00	1241.0
87.00	515.0
90.90	866.0
92.00	1148.0
93.00	6848.0
94.00	436.0
96.00	308.0
98.00	5311.0
99.00	4487.0
100.00	372.0
101.00	2355.0
103.00	914.0
104.00	1624.0
105.00	1480.0
106.00	535.0
107.00	19192.0
108.00	3004.0
109.10	707.0
110.00	37304.0
111.00	5522.0
112.00	684.0
115.90	1034.0
117.00	14884.0
118.00	1062.0
122.00	1462.0
123.00	2113.0
123.90	871.0
125.00	795.0
127.00	74424.0
128.00	5520.0
129.00	28720.0
130.00	2283.0
131.00	466.0
134.00	825.0
135.00	2276.0
136.00	920.0
137.00	1211.0
139.90	368.0
141.00	3773.0
142.00	1370.0
143.00	934.0
146.00	697.0
147.00	2018.0
148.00	4074.0
149.00	942.0
151.10	439.0
153.00	1258.0
154.00	966.0
155.00	2162.0
156.00	3183.0
157.00	703.0
157.90	733.0
158.90	469.0
160.00	1171.0
161.00	1735.0
162.00	550.0
165.00	1507.0
166.00	1236.0
167.00	7994.0
168.00	3981.0
169.00	694.0
170.90	378.0
171.90	722.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

173.00	918.0
174.00	1717.0
175.00	3185.0
176.00	864.0
176.90	1568.0
178.00	534.0
179.00	6143.0
180.00	4222.0
181.00	1991.0
182.00	340.0
183.90	443.0
185.00	2955.0
186.00	23856.0
187.00	6767.0
188.00	768.0
188.90	1371.0
191.00	653.0
192.00	1972.0
193.00	2204.0
194.00	536.0
196.00	5636.0
198.00	187392.0
199.00	12057.0
200.00	960.0
201.50	963.0
203.00	1178.0
204.00	6164.0
205.00	10878.0
206.00	47208.0
207.10	6264.0
208.00	1569.0
209.00	572.0
210.20	856.0
211.00	1850.0
214.90	456.0
216.00	1053.0
217.00	12317.0
218.00	1387.0
221.00	8688.0
223.00	2935.0
224.00	26032.0
225.00	6991.0
226.00	703.0
227.00	10568.0
228.00	1565.0
229.00	2349.0
230.00	346.0
231.10	949.0
234.00	643.0
235.00	786.0
236.00	513.0
237.00	978.0
238.90	457.0
240.00	322.0
240.90	603.0
242.00	1393.0
243.10	1684.0
244.00	22664.0
245.00	3017.0
246.00	4193.0
247.00	855.0
249.00	868.0
253.00	474.0
255.00	120528.0
256.00	17120.0
257.00	1413.0
258.00	6681.0
259.00	1046.0
265.00	2727.0
266.00	342.0
272.10	349.0
273.00	3948.0
274.00	10043.0
275.00	61376.0
276.00	8318.0
277.00	4761.0
278.00	786.0
282.90	451.0
283.90	365.0
285.00	899.0
293.00	1182.0
296.00	16202.0
297.00	2065.0
302.00	319.0
303.10	2011.0
304.00	550.0
314.00	786.0
315.00	1771.0
316.00	1153.0
321.00	564.0
322.00	329.0
323.10	5762.0
324.00	1092.0
327.00	1111.0
328.00	537.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

332.00	465.0
333.00	613.0
334.00	4022.0
335.00	1084.0
341.00	792.0
346.00	1260.0
352.00	1795.0
353.00	1300.0
354.00	2142.0
355.00	379.0
365.00	8711.0
366.00	1095.0
371.00	523.0
372.00	3561.0
373.00	885.0
383.00	1038.0
390.00	447.0
391.00	349.0
402.00	1394.0
403.00	2050.0
404.10	725.0
421.00	1688.0
422.10	1825.0
423.10	12093.0
424.00	2990.0
441.10	31072.0
442.10	197696.0
443.10	38040.0
444.10	3428.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

Scan 2481 (17.581 min): BN003729.D
DFTPP62

m/z	Abundance
44.00	841.0
207.00	330.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62