

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120720\
 Data File : BN013038.D
 Acq On : 10 Dec 2020 09:40
 Operator : CG/JU
 Sample : L4941-19
 Misc :
 ALS Vial : 108 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B52

Manual Integrations
 APPROVED

mohammad
 12/11/2020 1:18:54 PM

Quant Time: Dec 10 10:35:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 10 10:29:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	1195	0.40	ng/ul	0.00
2) Naphthalene-d8	10.14	136	4905	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.04	164	2665	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.80	188	5607	0.40	ng/ul	0.00
16) Chrysene-d12	21.03	240	4322	0.40	ng/ul	0.00
20) Perylene-d12	23.13	264	4167	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.75	152	1038	0.14	ng/ul	0.00
14) Fluoranthene-d10	18.85	212	2166	0.14	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	16.84	178	1354	0.072	ng/ul#	93
15) Fluoranthene	18.87	202	4151	0.186	ng/ul	98
17) Pyrene	19.24	202	3915	0.192	ng/ul	95
18) Benzo(a)anthracene	21.01	228	1706	0.095	ng/ul	93
19) Chrysene	21.07	228	2270	0.115	ng/ul	95
21) Benzo(b)fluoranthene	22.51	252	2790m	0.145	ng/ul	
22) Benzo(k)fluoranthene	22.55	252	1207m	0.057	ng/ul	
23) Benzo(a)pyrene	23.04	252	1862	0.104	ng/ul#	42
24) Indeno(1,2,3-cd)pyrene	25.18	276	1238	0.053	ng/ul#	98
26) Benzo(g,h,i)perylene	25.81	276	1128	0.053	ng/ul#	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120720\
Data File : BN013038.D
Acq On : 10 Dec 2020 09:40
Operator : CG/JU
Sample : L4941-19
Misc :
ALS Vial : 108 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B52

Manual Integrations
APPROVED

mohammad
12/11/2020 1:18:54 PM

Quant Time: Dec 10 10:35:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN120320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 10 10:29:27 2020
Response via : Initial Calibration

