


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120720\
 Data File : BN012931.D
 Acq On : 07 Dec 2020 09:29
 Operator : CG/JU
 Sample : DFTPP28
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP28

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN120320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 09 07:33:34 2020



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2507

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.3	133621	PASS
68	69	0.00	2	1.2	1564	PASS
69	198	0.00	100	50.0	127746	PASS
70	69	0.00	2	0.6	726	PASS
127	198	10	80	50.9	129917	PASS
197	198	0.00	2	0.4	941	PASS
198	198	100	100	100.0	255381	PASS
199	198	5	9	6.8	17368	PASS
275	198	10	60	28.3	72208	PASS
365	198	1	100	4.7	11930	PASS
441	198	0.01	100	13.9	35405	PASS
442	198	50	100	96.8	247168	PASS
443	442	15	24	18.5	45813	PASS

Scan 2507 (17.234 min): BN012931.D
DFTPP28

m/z	Abundance
39.00	453.0
44.00	1771.0
49.00	467.0
73.00	590.0
207.00	1163.0
208.00	406.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

m/z	Abundance
37.00	722.0
38.10	2341.0
39.10	11461.0
40.00	1182.0
41.00	459.0
44.00	1719.0
44.90	347.0
49.00	1728.0
50.00	39544.0
51.10	139840.0
52.10	6917.0
55.00	896.0
56.00	4558.0
57.00	8664.0
61.00	1884.0
62.00	1954.0
63.00	5458.0
64.00	590.0
65.00	1976.0
69.00	131328.0
70.10	590.0
73.00	1568.0
74.00	15660.0
75.00	22584.0
76.00	6694.0
77.10	142912.0
78.10	10739.0
79.00	12334.0
80.00	7985.0
81.00	11272.0
81.90	3071.0
83.10	1705.0
83.90	350.0
85.00	1527.0
86.00	3048.0
87.00	1686.0
88.00	421.0
91.00	2140.0
92.10	2658.0
93.00	18560.0
94.00	1116.0
95.90	764.0
98.00	16310.0
99.00	11817.0
100.10	761.0
101.00	4863.0
103.00	2026.0
104.00	4143.0
104.90	3630.0
106.00	1107.0
107.00	41216.0
108.00	6317.0
109.00	987.0
110.00	69064.0
111.00	12071.0
111.90	1094.0
113.10	465.0
116.10	3221.0
117.00	45632.0
118.00	3476.0
119.00	396.0
120.10	424.0
122.00	3313.0
123.00	3830.0
124.00	1705.0
125.10	1804.0
127.00	124744.0
128.00	9665.0
129.00	54768.0
130.00	5142.0
131.00	858.0
131.90	429.0
132.40	331.0
134.10	1428.0
135.00	4563.0
136.00	1543.0
137.00	3171.0
138.00	351.0
140.00	983.0
141.00	8891.0
142.10	2361.0
143.10	1792.0
144.00	533.0
145.00	466.0
146.00	1497.0
147.00	4318.0
148.00	12272.0
149.00	1840.0
150.00	340.0
150.90	907.0
151.80	770.0
153.00	2256.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

154.00	1298.0
155.00	4372.0
156.00	5589.0
157.10	1364.0
157.80	1355.0
158.90	928.0
159.90	2591.0
161.00	3252.0
162.00	837.0
162.90	340.0
165.00	2905.0
166.00	2448.0
167.00	17232.0
168.00	8698.0
169.00	1406.0
169.90	780.0
170.90	718.0
172.00	979.0
173.10	2011.0
174.00	3193.0
175.00	5367.0
176.10	1142.0
176.90	3341.0
178.10	1206.0
179.00	11715.0
180.00	7305.0
181.00	3034.0
182.10	682.0
184.00	1399.0
185.00	4882.0
186.00	34336.0
187.00	10341.0
188.00	1348.0
189.00	2561.0
189.90	470.0
190.90	832.0
192.00	3543.0
193.00	4096.0
194.00	735.0
195.10	423.0
196.00	5371.0
197.00	1291.0
198.00	224064.0
199.00	15048.0
200.00	1403.0
201.80	1095.0
203.00	2344.0
204.00	9842.0
205.00	16331.0
206.10	58280.0
207.10	9766.0
208.00	2439.0
208.90	1056.0
210.20	937.0
211.00	2442.0
212.00	373.0
215.00	745.0
216.00	1818.0
217.00	17536.0
218.00	2593.0
219.90	329.0
221.10	8255.0
222.00	2138.0
223.00	4545.0
224.00	37152.0
225.10	7880.0
226.10	893.0
227.00	15992.0
228.00	2530.0
228.90	2912.0
230.10	456.0
230.90	1205.0
234.00	852.0
235.00	1503.0
236.00	827.0
237.10	1490.0
239.00	621.0
240.00	528.0
241.10	832.0
242.00	1842.0
243.00	1789.0
244.00	23968.0
245.00	3197.0
246.00	6313.0
246.90	1586.0
249.00	750.0
250.90	518.0
252.00	377.0
253.00	602.0
255.00	133632.0
256.00	19552.0
257.00	1663.0
258.00	8699.0
258.90	1701.0
260.00	489.0
265.00	3841.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

266.00	781.0
270.90	435.0
272.10	360.0
273.00	3496.0
274.00	11072.0
275.00	57736.0
276.00	7899.0
277.00	6500.0
278.00	1184.0
281.00	323.0
281.90	346.0
283.10	944.0
284.00	851.0
285.00	641.0
289.00	332.0
290.00	401.0
291.80	369.0
292.90	1557.0
293.90	397.0
294.80	333.0
296.00	17704.0
297.00	2971.0
300.90	344.0
302.00	371.0
303.10	2612.0
304.10	665.0
308.00	311.0
309.90	481.0
314.00	894.0
315.00	2321.0
316.00	805.0
321.00	536.0
322.00	429.0
323.00	6027.0
324.00	1277.0
326.90	1589.0
327.90	537.0
332.00	479.0
333.00	526.0
333.90	3208.0
335.00	1054.0
341.10	1026.0
345.90	950.0
346.90	398.0
352.10	1532.0
353.00	1032.0
354.00	1508.0
354.90	372.0
365.00	8916.0
366.00	1315.0
370.90	654.0
372.00	2403.0
373.10	736.0
382.90	974.0
389.90	418.0
391.10	320.0
402.00	847.0
402.90	1323.0
403.80	649.0
421.00	920.0
422.00	1280.0
423.00	9165.0
424.00	2851.0
440.90	19424.0
442.00	143680.0
443.00	30072.0
444.00	3096.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

m/z	Abundance
37.00	792.0
38.10	2385.0
39.10	12169.0
40.00	1073.0
41.10	302.0
44.00	1864.0
49.00	1569.0
50.10	46912.0
51.10	162304.0
52.10	8565.0
55.00	1094.0
56.00	4879.0
57.00	9951.0
57.90	494.0
60.90	1572.0
62.00	3223.0
63.00	5371.0
64.00	1016.0
65.00	2287.0
68.00	2855.0
69.00	156672.0
69.90	827.0
73.10	1543.0
74.00	20920.0
75.00	25512.0
76.10	9020.0
77.00	172608.0
78.10	12221.0
79.00	13692.0
80.00	9602.0
81.00	13026.0
82.00	3578.0
83.00	2780.0
83.70	464.0
85.00	2307.0
86.00	4466.0
87.10	1629.0
88.20	365.0
91.00	3400.0
92.00	3890.0
93.00	25304.0
94.00	1547.0
95.00	323.0
96.00	617.0
98.00	19912.0
99.00	14293.0
100.00	794.0
101.00	7301.0
102.10	424.0
103.00	1937.0
104.00	5256.0
105.00	4735.0
106.10	1627.0
107.00	53144.0
108.00	7411.0
109.10	1788.0
110.00	88608.0
111.00	15146.0
112.00	2103.0
112.90	383.0
114.90	366.0
116.00	3075.0
117.00	64928.0
118.00	4438.0
119.00	721.0
119.90	514.0
122.00	4415.0
123.00	6250.0
124.00	2572.0
125.00	2217.0
126.10	912.0
127.00	163072.0
128.00	11250.0
129.00	66968.0
130.00	6745.0
131.00	1303.0
132.00	507.0
132.80	525.0
134.00	2617.0
135.00	5998.0
136.00	2039.0
137.00	3555.0
137.90	663.0
139.00	458.0
140.00	902.0
141.00	10358.0
142.00	3201.0
143.00	1936.0
144.10	474.0
144.90	558.0
146.00	2422.0
147.00	5861.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

148.00	13451.0
149.00	2958.0
149.90	580.0
151.10	805.0
151.60	578.0
152.00	513.0
153.00	3174.0
154.00	2231.0
155.10	5488.0
156.00	8297.0
156.90	1284.0
157.90	1701.0
158.90	1336.0
160.00	3177.0
161.00	4788.0
162.10	1294.0
163.00	332.0
163.80	436.0
165.00	5019.0
166.10	2931.0
167.00	21008.0
168.00	9816.0
169.10	1820.0
170.00	993.0
171.00	569.0
171.90	1728.0
173.00	2251.0
174.00	3557.0
175.00	7490.0
176.00	1379.0
176.20	1401.0
176.90	4325.0
178.00	1293.0
179.00	13849.0
180.00	10284.0
181.00	4024.0
181.90	699.0
183.90	1259.0
185.00	6175.0
186.00	46408.0
187.00	13628.0
188.00	1469.0
189.00	3929.0
190.00	793.0
191.00	1458.0
192.00	4505.0
193.00	5489.0
194.10	1017.0
195.00	501.0
196.10	7999.0
198.00	311360.0
199.00	22520.0
200.10	1945.0
201.60	1504.0
203.10	3141.0
204.00	13667.0
205.00	21856.0
206.10	88400.0
207.00	10908.0
208.00	3960.0
208.90	1124.0
210.20	1748.0
211.00	3488.0
212.00	550.0
214.90	1546.0
216.00	2257.0
217.00	26280.0
218.10	3505.0
219.00	340.0
221.10	10679.0
221.90	2780.0
223.00	5185.0
224.00	46848.0
225.00	11690.0
226.00	1985.0
227.00	23880.0
228.10	3381.0
229.00	4437.0
230.00	535.0
231.10	1351.0
232.10	401.0
233.10	382.0
234.00	1994.0
235.00	1286.0
235.90	1070.0
237.00	2315.0
237.90	529.0
238.90	880.0
240.00	722.0
240.90	1308.0
242.00	2789.0
243.00	2575.0
244.00	38936.0
245.10	5118.0
246.00	9093.0
247.10	1832.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

249.00	1507.0
251.90	474.0
253.10	1024.0
255.00	196544.0
256.00	28736.0
257.00	2517.0
258.00	13047.0
259.00	1998.0
259.90	434.0
263.90	433.0
265.00	5617.0
266.00	1030.0
270.90	582.0
272.10	541.0
273.00	6480.0
274.10	15185.0
275.00	86752.0
276.00	11593.0
277.00	7599.0
278.10	1552.0
279.00	594.0
281.00	423.0
282.20	438.0
283.00	1288.0
284.00	771.0
285.10	1709.0
285.90	587.0
292.00	337.0
293.00	1684.0
293.90	348.0
296.00	30168.0
297.00	4332.0
301.00	464.0
301.80	551.0
303.00	3093.0
303.90	1038.0
308.00	415.0
309.00	488.0
309.90	663.0
314.00	1322.0
315.00	3012.0
316.10	1763.0
317.00	376.0
321.00	858.0
321.90	341.0
323.00	7805.0
324.10	1601.0
327.00	1822.0
327.90	1258.0
328.90	376.0
332.00	655.0
333.00	1352.0
334.00	4843.0
335.00	1425.0
341.10	1452.0
345.90	2442.0
347.00	400.0
352.00	3369.0
353.00	1576.0
353.90	2265.0
355.00	694.0
365.00	14573.0
366.00	1372.0
370.10	583.0
371.00	674.0
372.00	4643.0
373.00	998.0
383.00	1216.0
383.80	426.0
389.90	767.0
401.00	459.0
402.00	2216.0
402.90	2109.0
403.90	1058.0
421.00	2650.0
422.10	1646.0
423.00	13912.0
424.00	4188.0
424.80	568.0
441.00	45144.0
442.00	301440.0
443.00	53368.0
444.10	5830.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

m/z	Abundance
37.00	613.0
38.10	2079.0
39.10	7004.0
40.00	576.0
44.00	2095.0
49.20	1373.0
50.10	28192.0
51.10	98720.0
52.10	5279.0
55.00	532.0
56.00	2928.0
57.00	5576.0
57.90	455.0
61.00	1442.0
62.10	2065.0
63.00	3714.0
64.00	602.0
65.00	1709.0
68.10	1838.0
69.00	95240.0
70.00	763.0
73.10	1526.0
74.00	10694.0
75.00	15610.0
76.10	4990.0
77.10	111120.0
78.00	7258.0
79.00	8763.0
80.00	5816.0
81.00	7071.0
81.90	2165.0
83.00	1647.0
84.10	497.0
85.00	1116.0
86.00	3297.0
86.90	1173.0
88.00	344.0
88.90	332.0
91.00	1788.0
91.90	2620.0
93.00	16424.0
93.90	1283.0
96.10	565.0
98.00	13325.0
99.00	8854.0
99.90	785.0
101.00	4740.0
101.90	385.0
103.00	1811.0
104.00	3297.0
105.00	2412.0
105.90	889.0
107.00	34608.0
108.00	4327.0
110.00	57096.0
111.00	8933.0
111.90	1400.0
113.00	462.0
115.90	1882.0
117.00	38472.0
117.90	2404.0
118.90	486.0
121.90	2296.0
123.00	4256.0
124.00	2257.0
125.00	1884.0
127.00	101936.0
128.00	9093.0
129.00	44336.0
129.90	4410.0
131.00	903.0
131.90	548.0
133.00	363.0
134.00	1616.0
135.00	3572.0
136.10	1189.0
137.00	2577.0
137.80	419.0
139.00	496.0
139.80	461.0
140.00	444.0
140.90	6200.0
141.90	2074.0
143.00	1332.0
145.10	354.0
146.00	1001.0
147.00	4098.0
148.00	9289.0
148.90	2135.0
150.10	643.0
151.00	1033.0
151.80	532.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

153.00	2420.0
154.00	1395.0
154.90	3595.0
156.00	4711.0
157.10	728.0
157.90	1143.0
159.00	843.0
160.00	2236.0
161.00	2818.0
162.10	592.0
162.90	331.0
164.00	425.0
164.90	2699.0
165.90	2141.0
167.00	16022.0
168.00	7635.0
169.00	1006.0
169.90	317.0
171.00	605.0
171.90	1455.0
173.10	1310.0
174.00	2658.0
175.00	5541.0
176.10	853.0
177.00	3105.0
178.00	765.0
179.00	10783.0
180.00	7086.0
180.90	2675.0
181.90	423.0
184.00	1576.0
185.00	5075.0
186.00	34872.0
187.00	10012.0
188.00	882.0
189.00	2571.0
190.00	608.0
191.10	1418.0
192.00	3247.0
193.00	3599.0
194.00	720.0
194.90	805.0
196.10	6949.0
196.90	1533.0
198.00	230720.0
199.00	14538.0
200.00	1099.0
201.40	1259.0
203.10	2163.0
204.10	9837.0
205.00	17456.0
206.10	64040.0
207.00	9285.0
207.90	2874.0
208.90	1022.0
210.00	1070.0
211.00	3066.0
212.00	364.0
215.00	1069.0
216.00	1562.0
217.00	19592.0
218.00	2334.0
220.10	567.0
221.00	8847.0
222.00	2630.0
223.10	3650.0
224.00	37464.0
225.10	9739.0
226.00	630.0
227.00	18920.0
228.00	2356.0
229.00	3257.0
229.90	687.0
230.90	1279.0
234.00	1211.0
234.90	1397.0
236.10	401.0
237.00	1479.0
239.00	704.0
239.90	694.0
241.00	1045.0
242.10	1943.0
243.10	2509.0
244.00	30544.0
245.10	4273.0
246.00	7405.0
246.90	1325.0
248.00	385.0
249.00	677.0
251.90	745.0
253.00	821.0
255.00	157824.0
256.00	22840.0
257.10	1579.0
258.00	11116.0
259.00	1961.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28

265.00	5005.0
266.00	400.0
269.90	308.0
270.80	387.0
272.00	891.0
273.00	4927.0
274.00	13418.0
275.00	72136.0
276.00	7992.0
277.00	6981.0
277.90	1725.0
280.90	372.0
282.00	309.0
282.90	878.0
284.00	698.0
285.00	582.0
285.90	379.0
288.90	408.0
290.90	370.0
292.10	426.0
292.90	1414.0
293.80	467.0
296.00	25512.0
297.00	3684.0
301.00	319.0
302.00	535.0
303.00	2361.0
303.90	831.0
307.90	350.0
310.00	325.0
314.00	926.0
315.00	3839.0
315.90	1358.0
321.00	445.0
323.00	7471.0
324.00	1505.0
326.90	1615.0
328.00	645.0
331.90	549.0
333.00	1108.0
334.00	5011.0
335.00	1609.0
341.00	1190.0
345.90	1797.0
347.00	449.0
352.00	2159.0
353.00	1628.0
354.00	2001.0
354.90	439.0
364.10	364.0
365.00	12301.0
366.00	1664.0
370.10	337.0
371.00	706.0
372.00	3974.0
373.00	1074.0
383.00	799.0
389.80	578.0
400.90	486.0
402.00	2329.0
403.10	2644.0
403.90	819.0
421.00	2286.0
422.00	1961.0
423.00	13780.0
424.00	3369.0
424.90	719.0
440.10	415.0
441.10	41648.0
442.00	296384.0
443.10	54000.0
444.10	4415.0
444.90	305.0

Instrument :
BNA_N
ClientSampleId :
DFTPP28