

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
 Data File : BN029126.D
 Acq On : 08 Dec 2023 14:11
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Dec 08 15:21:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 00:49:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.789	152	9	0.400	ng	# 0.01
7) Naphthalene-d8	10.573	136	10	0.400	ng	# 0.01
13) Acenaphthene-d10	14.300	164	11	0.400	ng	#-0.10
19) Phenanthrene-d10	17.072	188	71	0.400	ng	#-0.07
29) Chrysene-d12	21.311	240	22	0.400	ng	#-0.04
35) Perylene-d12	23.620	264	1	0.400	ng	#-0.05
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	2	0.098	ng	0.00
5) Phenol-d6	6.995	99	3	0.123	ng	0.01
8) Nitrobenzene-d5	8.971	82	20	1.658	ng	0.03
11) 2-Methylnaphthalene-d10	12.004	152	4	0.190	ng	-0.14
14) 2,4,6-Tribromophenol	15.756	330	1	0.082	ng	-0.14
15) 2-Fluorobiphenyl	12.942	172	1	0.018	ng	-0.09
27) Fluoranthene-d10	19.102	212	3	0.013	ng	-0.08
31) Terphenyl-d14	19.752	244	2	0.036	ng	-0.04
Target Compounds						
2) 1,4-Dioxane	3.369	88	6	0.828	ng	# 41
3) n-Nitrosodimethylamine	3.745	42	2	0.455	ng	# 1
6) bis(2-Chloroethyl)ether	7.255	93	5	0.290	ng	# 1
9) Naphthalene	10.648	128	6	0.192	ng	# 1
10) Hexachlorobutadiene	10.925	225	3	0.218	ng	# 19
12) 2-Methylnaphthalene	12.253	142	2	0.079	ng	# 49
16) Acenaphthylene	14.033	152	103	1.701	ng	# 1
17) Acenaphthene	14.343	154	4	0.105	ng	# 3
18) Fluorene	15.326	166	4	0.066	ng	# 8
20) 4,6-Dinitro-2-methylph...	15.497	198	4	0.187	ng	# 1
22) Hexachlorobenzene	16.377	284	4	0.055	ng	# 40
23) Atrazine	16.551	200	4	0.071	ng	# 1
24) Pentachlorophenol	16.712	266	14	0.350	ng	# 37
25) Phenanthrene	17.072	178	32	0.135	ng	# 1
26) Anthracene	17.184	178	18	0.079	ng	# 80
32) Benzo(a)anthracene	21.293	228	7	0.063	ng	# 1
33) Chrysene	21.365	228	7	0.064	ng	# 1
34) Bis(2-ethylhexyl)phtha...	21.275	149	63	0.916	ng	# 83
36) Indeno(1,2,3-cd)pyrene	25.982	276	2	0.392	ng	# 1
37) Benzo(b)fluoranthene	22.924	252	9	1.676	ng	# 1
38) Benzo(k)fluoranthene	22.965	252	15	2.980	ng	# 1
39) Benzo(a)pyrene	23.520	252	9	2.026	ng	# 1
40) Dibenzo(a,h)anthracene	26.008	278	4	1.017	ng	# 1
41) Benzo(g,h,i)perylene	26.722	276	1	0.232	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
Data File : BN029126.D
Acq On : 08 Dec 2023 14:11
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Dec 08 15:21:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 06 00:49:47 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.789 min Scan# 61
 Delta R.T. 0.014 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 9
 Ion Ratio Lower Upper
 152 100
 150 92.5 116.2 174.4#
 115 134.0 45.1 67.7#



#2
 1,4-Dioxane
 Concen: 0.828 ng
 RT: 3.369 min Scan# 40
 Delta R.T. 0.029 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11

Tgt Ion: 88 Resp: 6
 Ion Ratio Lower Upper
 88 100
 43 0.0 15.6 23.4#
 58 0.0 33.2 49.8#









#3
 n-Nitrosodimethylamine
 Concen: 0.455 ng
 RT: 3.745 min Scan# 91
 Delta R.T. 0.022 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 42 Resp: 2
 Ion Ratio Lower Upper
 42 100
 74 750.0 106.3 159.5#
 44 250.0 0.0 0.0#



#4
 2-Fluorophenol
 Concen: 0.098 ng
 RT: 5.392 min Scan# 320
 Delta R.T. 0.000 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11

Tgt Ion: 112 Resp: 2
 Ion Ratio Lower Upper
 112 100
 64 0.0 37.0 55.6#
 63 250.0 29.0 43.6#









#5
Phenol-d6
Concen: 0.123 ng
RT: 6.995 min Scan# 541
Delta R.T. 0.014 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 3
Ion Ratio Lower Upper
99 100
42 300.0 13.4 20.2#
71 100.0 47.8 71.6#



#6
bis(2-Chloroethyl)ether
Concen: 0.290 ng
RT: 7.255 min Scan# 578
Delta R.T. 0.022 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion: 93 Resp: 5
Ion Ratio Lower Upper
93 100
63 140.0 48.4 72.6#
95 140.0 28.1 42.1#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.573 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136 Resp: 10
Ion Ratio Lower Upper
136 100
137 100.0 11.9 17.9#
54 100.0 6.4 9.6#
68 103.9 5.3 7.9#



#8
Nitrobenzene-d5
Concen: 1.658 ng
RT: 8.971 min Scan# 786
Delta R.T. 0.032 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion: 82 Resp: 20
Ion Ratio Lower Upper
82 100
128 82.3 28.2 42.4#
54 83.9 32.1 48.1#









#9
Naphthalene
Concen: 0.192 ng
RT: 10.648 min Scan# 941
Delta R.T. 0.043 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 6
Ion Ratio Lower Upper
128 100
129 88.9 10.5 15.7#
127 88.9 12.5 18.7#



#10
Hexachlorobutadiene
Concen: 0.218 ng
RT: 10.925 min Scan# 969
Delta R.T. 0.043 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:225 Resp: 3
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 0.0 50.7 76.1#









#11
2-Methylnaphthalene-d10
Concen: 0.190 ng
RT: 12.004 min Scan# 1192
Delta R.T. -0.144 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 4
Ion Ratio Lower Upper
152 100
151 25.0 17.4 26.0



#12
2-Methylnaphthalene
Concen: 0.079 ng
RT: 12.253 min Scan# 1220
Delta R.T. 0.030 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:142 Resp: 2
Ion Ratio Lower Upper
142 100
141 102.1 71.1 106.7
115 136.2 44.6 66.8#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.300 min Scan# 14
Delta R.T. -0.096 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 11
Ion Ratio Lower Upper
164 100
162 94.3 85.0 127.6
160 86.8 41.3 61.9#



#14
2,4,6-Tribromophenol
Concen: 0.082 ng
RT: 15.756 min Scan# 1622
Delta R.T. -0.136 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:330 Resp: 1
Ion Ratio Lower Upper
330 100
332 0.0 78.8 118.2#
141 300.0 28.7 43.1#









#15
2-Fluorobiphenyl
Concen: 0.018 ng
RT: 12.942 min Scan# 11
Delta R.T. -0.085 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 1
Ion Ratio Lower Upper
172 100
171 100.0 30.1 45.1#
170 97.8 20.3 30.5#



#16
Acenaphthylene
Concen: 1.701 ng
RT: 14.033 min Scan# 1464
Delta R.T. -0.085 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:152 Resp: 103
Ion Ratio Lower Upper
152 100
151 756.3 15.8 23.6#
153 12.6 10.3 15.5









#17
Acenaphthene
Concen: 0.105 ng
RT: 14.343 min Scan# 14
Delta R.T. -0.118 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 4
Ion Ratio Lower Upper
154 100
153 0.0 93.0 139.6#
152 0.0 44.2 66.4#



#18
Fluorene
Concen: 0.066 ng
RT: 15.326 min Scan# 1585
Delta R.T. -0.128 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:166 Resp: 4
Ion Ratio Lower Upper
166 100
165 0.0 80.2 120.4#
167 0.0 11.0 16.4#





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.072 min Scan# 11

Delta R.T. -0.074 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:188 Resp: 71

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 68.5 5.4 8.0#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.187 ng

RT: 15.497 min Scan# 1601

Delta R.T. -0.048 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11




Tgt Ion:198 Resp: 4


Ion Ratio Lower Upper

198 100

51 161.7 23.4 35.2#

105 148.9 29.0 43.4#









#22
Hexachlorobenzene
Concen: 0.055 ng
RT: 16.377 min Scan# 1685
Delta R.T. -0.062 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:284 Resp: 4
Ion Ratio Lower Upper
284 100
142 0.0 28.0 42.0#
249 0.0 27.8 41.8#



#23
Atrazine
Concen: 0.071 ng
RT: 16.551 min Scan# 1686
Delta R.T. -0.087 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:200 Resp: 4
Ion Ratio Lower Upper
200 100
173 119.1 19.3 28.9#
215 102.1 36.1 54.1#





#24

Pentachlorophenol

Concen: 0.350 ng

RT: 16.712 min Scan# 1

Delta R.T. -0.087 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:266 Resp: 14

Ion Ratio Lower Upper

266 100

264 0.0 50.7 76.1#

268 28.6 49.5 74.3#



#25

Phenanthrene

Concen: 0.135 ng

RT: 17.072 min Scan# 1728

Delta R.T. -0.112 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11




Tgt Ion:178 Resp: 32


Ion Ratio Lower Upper

178 100

176 0.0 15.6 23.4#

179 884.4 12.6 18.8#









#26
 Anthracene
 Concen: 0.079 ng
 RT: 17.184 min Scan# 11
 Delta R.T. -0.099 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:178 Resp: 18
 Ion Ratio Lower Upper
 178 100
 176 16.7 15.0 22.4
 179 0.0 12.6 19.0#



#27
 Fluoranthene-d10
 Concen: 0.013 ng
 RT: 19.102 min Scan# 1971
 Delta R.T. -0.084 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11

Tgt Ion:212 Resp: 3
 Ion Ratio Lower Upper
 212 100
 106 0.0 3.9 5.9#
 104 33.3 2.6 3.8#





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.311 min Scan# 2419

Delta R.T. -0.036 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 22

Ion Ratio Lower Upper

240 100

120 146.8 6.7 10.1#

236 108.1 25.7 38.5#



#31

Terphenyl-d14

Concen: 0.036 ng

RT: 19.752 min Scan# 2111

Delta R.T. -0.042 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11




Tgt Ion:244 Resp: 2


Ion Ratio Lower Upper

244 100

212 3289.6 11.6 17.4#

122 162.5 8.2 12.4#









#32
Benzo(a)anthracene
Concen: 0.063 ng
RT: 21.293 min Scan# 2413
Delta R.T. -0.036 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228 Resp: 7
Ion Ratio Lower Upper
228 100
226 105.3 23.8 35.6#
229 103.5 16.0 24.0#



#33
Chrysene
Concen: 0.064 ng
RT: 21.365 min Scan# 2421
Delta R.T. -0.018 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:228 Resp: 7
Ion Ratio Lower Upper
228 100
226 116.4 25.4 38.0#
229 107.3 16.0 24.0#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.916 ng
 RT: 21.275 min Scan# 2411
 Delta R.T. 0.000 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	149	Resp:	63
Ion	Ratio	Lower	Upper
149	100		
167	19.0	23.4	35.2#
279	4.8	3.4	5.2



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.620 min Scan# 2953
 Delta R.T. -0.050 min
 Lab File: BN029126.D
 Acq: 08 Dec 2023 14:11

Tgt Ion:	264	Resp:	1
Ion	Ratio	Lower	Upper
264	100		
260	90.2	23.9	35.9#
265	870.5	52.7	79.1#





#36

Indeno(1,2,3-cd)pyrene

Concen: 0.392 ng

RT: 25.982 min Scan# 31

Delta R.T. -0.058 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:276 Resp: 2

Ion Ratio Lower Upper

276 100

138 150.0 9.6 14.4#

277 0.0 19.4 29.0#



#37

Benzo(b)fluoranthene

Concen: 1.676 ng

RT: 22.924 min Scan# 2715

Delta R.T. -0.041 min

Lab File: BN029126.D

Acq: 08 Dec 2023 14:11




Tgt Ion:252 Resp: 9


Ion Ratio Lower Upper

252 100

253 167.9 23.0 34.6#


125 65.5 5.4 8.0#








#38
Benzo(k)fluoranthene
Concen: 2.980 ng
RT: 22.965 min Scan# 21
Delta R.T. -0.047 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 15
Ion Ratio Lower Upper
252 100
253 171.1 23.0 34.4#
125 66.3 5.0 7.6#



#39
Benzo(a)pyrene
Concen: 2.026 ng
RT: 23.520 min Scan# 2919
Delta R.T. -0.044 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:252 Resp: 9
Ion Ratio Lower Upper
252 100
253 171.1 25.3 37.9#
125 71.7 6.5 9.7#









#40
Dibenzo(a,h)anthracene
Concen: 1.017 ng
RT: 26.008 min Scan# 31
Delta R.T. -0.064 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:278 Resp: 4
Ion Ratio Lower Upper
278 100
139 133.9 9.4 14.2#
279 216.1 24.2 36.4#



#41
Benzo(g,h,i)perylene
Concen: 0.232 ng
RT: 26.722 min Scan# 4014
Delta R.T. -0.047 min
Lab File: BN029126.D
Acq: 08 Dec 2023 14:11

Tgt Ion:276 Resp: 1
Ion Ratio Lower Upper
276 100
277 106.9 20.3 30.5#
138 120.7 9.2 13.8#

