

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
 Data File : BN029106.D
 Acq On : 08 Dec 2023 00:19
 Operator : MA/JU
 Sample : 05404-23DL 2X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D2-20231109DL

Quant Time: Dec 08 01:04:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 00:49:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.768	152	952	0.400	ng	0.00
7) Naphthalene-d8	10.551	136	1863	0.400	ng	-0.01
13) Acenaphthene-d10	14.385	164	1291	0.400	ng	-0.01
19) Phenanthrene-d10	17.146	188	2602	0.400	ng	# 0.00
29) Chrysene-d12	21.338	240	1481	0.400	ng	# 0.00
35) Perylene-d12	23.655	264	1632	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.377	112	122	0.057	ng	-0.01
5) Phenol-d6	6.973	99	74	0.029	ng	0.00
8) Nitrobenzene-d5	8.939	82	304	0.135	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	419	0.107	ng	-0.01
14) 2,4,6-Tribromophenol	15.893	330	53	0.037	ng	0.00
15) 2-Fluorobiphenyl	13.028	172	989	0.148	ng	0.00
27) Fluoranthene-d10	19.176	212	1061	0.130	ng	0.00
31) Terphenyl-d14	19.785	244	687	0.183	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.319	88	3197	4.170	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
Data File : BN029106.D
Acq On : 08 Dec 2023 00:19
Operator : MA/JU
Sample : 05404-23DL 2X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D2-20231109DL

Quant Time: Dec 08 01:04:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 06 00:49:47 2023
Response via : Initial Calibration

