


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
Data File : BN029113.D
Acq On : 08 Dec 2023 05:41
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120523.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Dec 06 00:49:47 2023






AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.6	38312	PASS
68	69	0.00	2	1.9	1308	PASS
69	198	0.00	100	54.9	68651	PASS
70	69	0.00	2	0.7	448	PASS
127	198	10	80	48.3	60429	PASS
197	198	0.00	2	0.8	956	PASS
198	198	100	100	100.0	125027	PASS
199	198	5	9	7.0	8804	PASS
275	198	10	60	27.5	34384	PASS
365	198	1	100	6.4	8010	PASS
441	198	0.01	100	11.8	14778	PASS
442	442	50	100	100.0	85909	PASS
443	442	15	24	19.7	16899	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
Data File : BN029113.D
Acq On : 08 Dec 2023 05:41
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 08 06:13:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:28:24 2023
Response via : Initial Calibration



TIC: BN029113.D\data.ms

(70) Pentachlorophenol (C)

16.793min (-0.023) 12735.60 ng




response 117936

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	55.06
264.00	61.60	60.40
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120723\
Data File : BN029113.D
Acq On : 08 Dec 2023 05:41
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 08 06:13:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 23:28:24 2023
Response via : Initial Calibration



TIC: BN029113.D\data.ms

(77) Benzidine

19.422min (+ 0.018) 0.00 ng

response 559605

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.02
183.00	13.20	11.04
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/7/2023	BNA_N	<u>BN029113.D</u>
Compound Name	Response	Retention Time
DDT	431537	20.639
DDD	10098	20.198
DDE	383	19.686
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10481	442018	2.37

Instrument :
BNA_N
ClientSampleId :
DFTPP