

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120919\
 Data File : BN008896.D
 Acq On : 09 Dec 2019 17:49
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN120919

Quant Time: Dec 09 18:26:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.375	0.406	-8.3	89	0.00
3	n-Nitrosodimethylamine	0.518	0.443	14.5	86	0.00
4 S	2-Fluorophenol	1.017	1.022	-0.5	91	0.00
5 S	Phenol-d6	1.406	1.393	0.9	91	0.00
6	bis(2-Chloroethyl)ether	1.325	1.381	-4.2	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.339	0.328	3.2	92	0.00
9	Naphthalene	1.091	1.151	-5.5	93	0.00
10	Hexachlorobutadiene	0.209	0.222	-6.2	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.684	0.717	-4.8	92	0.00
12	2-Methylnaphthalene	0.813	0.862	-6.0	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.192	0.181	5.7	88	0.00
15 S	2-Fluorobiphenyl	1.519	1.549	-2.0	92	0.00
16	Acenaphthylene	1.830	1.911	-4.4	91	0.00
17	Acenaphthene	1.215	1.291	-6.3	91	0.00
18	Fluorene	1.634	1.745	-6.8	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.239	0.242	-1.3	91	0.00
21	Hexachlorobenzene	0.273	0.277	-1.5	90	0.00
22	Pentachlorophenol	0.098	0.087	11.2	88	0.00
23	Phenanthrene	1.259	1.277	-1.4	90	0.00
24	Anthracene	1.109	1.100	0.8	90	0.00
25 SURR	Fluoranthene-d10	1.211	1.195	1.3	90	0.00
26	Fluoranthene	1.498	1.509	-0.7	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
28	Pvrene	1.402	1.506	-7.4	90	0.00
29 S	Terphenyl-d14	0.913	0.935	-2.4	90	0.00
30	Benzo(a)anthracene	1.418	1.496	-5.5	90	0.00
31	Chrysene	1.470	1.577	-7.3	89	-0.01
32	Bis(2-ethylhexyl)phthalate	0.664	0.576	13.3	84	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.649	-6.0	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.432	1.493	-4.3	91	0.00
36	Benzo(k)fluoranthene	1.442	1.576	-9.3	97	0.00
37 C	Benzo(a)pyrene	1.286	1.345	-4.6	94	0.00
38	Dibenzo(a,h)anthracene	1.206	1.279	-6.1	99	0.00
39	Benzo(a,h,i)perylene	1.191	1.261	-5.9	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN120919\
Data File : BN008896.D
Acq On : 09 Dec 2019 17:49
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN120919

Quant Time: Dec 09 18:26:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:02:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			