

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121019\
 Data File : BN008898.D
 Acq On : 10 Dec 2019 10:14
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 10 15:34:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.375	0.396	-5.6	102	0.00
3	n-Nitrosodimethylamine	0.518	0.453	12.5	103	0.00
4 S	2-Fluorophenol	1.017	1.028	-1.1	107	0.00
5 S	Phenol-d6	1.406	1.398	0.6	107	0.00
6	bis(2-Chloroethyl)ether	1.325	1.401	-5.7	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.339	0.327	3.5	107	0.00
9	Naphthalene	1.091	1.150	-5.4	108	0.00
10	Hexachlorobutadiene	0.209	0.218	-4.3	108	-0.01
11 SURR	2-Methylnaphthalene-d10	0.684	0.719	-5.1	109	0.00
12	2-Methylnaphthalene	0.813	0.866	-6.5	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.192	0.186	3.1	106	0.00
15 S	2-Fluorobiphenyl	1.519	1.541	-1.4	109	-0.01
16	Acenaphthylene	1.830	1.912	-4.5	107	0.00
17	Acenaphthene	1.215	1.286	-5.8	107	0.00
18	Fluorene	1.634	1.765	-8.0	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.239	0.240	-0.4	107	0.00
21	Hexachlorobenzene	0.273	0.277	-1.5	108	0.00
22	Pentachlorophenol	0.098	0.094	4.1	114	0.00
23	Phenanthrene	1.259	1.286	-2.1	108	0.00
24	Anthracene	1.109	1.105	0.4	107	0.00
25 SURR	Fluoranthene-d10	1.211	1.199	1.0	108	0.00
26	Fluoranthene	1.498	1.520	-1.5	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
28	Pvrene	1.402	1.500	-7.0	108	0.00
29 S	Terphenyl-d14	0.913	0.935	-2.4	109	0.00
30	Benzo(a)anthracene	1.418	1.498	-5.6	109	0.00
31	Chrysene	1.470	1.548	-5.3	106	-0.01
32	Bis(2-ethylhexyl)phthalate	0.664	0.636	4.2	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.590	-2.2	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.432	1.497	-4.5	107	0.00
36	Benzo(k)fluoranthene	1.442	1.578	-9.4	114	0.00
37 C	Benzo(a)pyrene	1.286	1.355	-5.4	111	0.00
38	Dibenzo(a,h)anthracene	1.206	1.260	-4.5	114	0.00
39	Benzo(a,h,i)perylene	1.191	1.237	-3.9	112	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121019\
Data File : BN008898.D
Acq On : 10 Dec 2019 10:14
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 10 15:34:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:02:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			