

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121020\
 Data File : BN013060.D
 Acq On : 11 Dec 2020 01:20
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN121020

Quant Time: Dec 11 01:56:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 11 01:53:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.949	1.017	-7.2	97	0.00
3	n-Nitrosodimethylamine	1.050	1.134	-8.0	127	0.00
4 S	2-Fluorophenol	1.437	1.490	-3.7	116	0.00
5 S	Phenol-d6	1.646	1.651	-0.3	115	0.00
6	bis(2-Chloroethyl)ether	1.111	0.994	10.5	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.424	0.394	7.1	112	0.00
9	Naphthalene	1.224	1.190	2.8	116	0.00
10	Hexachlorobutadiene	0.265	0.267	-0.8	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.722	0.698	3.3	114	0.00
12	2-Methylnaphthalene	0.839	0.785	6.4	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.263	0.225	14.4	103	0.00
15 S	2-Fluorobiphenyl	1.762	1.685	4.4	111	0.00
16	Acenaphthylene	2.132	2.019	5.3	113	0.00
17	Acenaphthene	1.345	1.316	2.2	112	0.00
18	Fluorene	1.795	1.682	6.3	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.071	6.6	107	0.00
21	4-Bromophenyl-phenylether	0.093	0.079	15.1	57	0.01
22	Hexachlorobenzene	0.294	0.279	5.1	109	0.00
23	Atrazine	0.238	0.211	11.3	108	0.00
24	Pentachlorophenol	0.123	0.102	17.1	106	0.01
25	Phenanthrene	1.314	1.225	6.8	110	0.00
26	Anthracene	1.167	1.042	10.7	109	0.00
27 SURR	Fluoranthene-d10	1.343	1.251	6.9	110	0.00
28	Fluoranthene	1.621	1.489	8.1	109	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
30	Pyrene	1.500	1.399	6.7	108	0.00
31 S	Terphenyl-d14	0.998	0.943	5.5	109	0.00
32	Benzo(a)anthracene	1.466	1.328	9.4	101	0.00
33	Chrysene	1.524	1.491	2.2	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.893	0.794	11.1	102	0.00
35	Indeno(1,2,3-cd)pyrene	2.024	1.900	6.1	103	0.00
36 I	Perylene-d12	1.000	1.000	0.0	98	0.00
37	Benzo(b)fluoranthene	1.520	1.422	6.4	108	0.00
38	Benzo(k)fluoranthene	1.605	1.551	3.4	108	0.00
39 C	Benzo(a)pyrene	1.450	1.391	4.1	104	0.00
40	Dibenzo(a,h)anthracene	1.582	1.514	4.3	105	0.00
41	Benzo(g,h,i)perylene	1.673	1.627	2.7	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN121020