

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121020\
 Data File : BN013060.D
 Acq On : 11 Dec 2020 01:20
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN121020

Quant Time: Dec 11 01:56:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 11 01:53:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.460	-15.0	97	0.00
3	n-Nitrosodimethylamine	0.400	0.432	-8.0	127	0.00
4 S	2-Fluorophenol	0.400	0.415	-3.7	116	0.00
5 S	Phenol-d6	0.400	0.401	-0.3	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.371	7.3	112	0.00
9	Naphthalene	0.400	0.389	2.8	116	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	114	0.00
12	2-Methylnaphthalene	0.400	0.374	6.5	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	111	0.00
16	Acenaphthylene	0.400	0.379	5.3	113	0.00
17	Acenaphthene	0.400	0.391	2.3	112	0.00
18	Fluorene	0.400	0.375	6.3	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.375	6.3	107	0.00
21	4-Bromophenyl-phenylether	0.400	0.340	15.0	57	0.01
22	Hexachlorobenzene	0.400	0.381	4.8	109	0.00
23	Atrazine	0.400	0.355	11.3	108	0.00
24	Pentachlorophenol	0.400	0.332	17.0	106	0.01
25	Phenanthrene	0.400	0.373	6.8	110	0.00
26	Anthracene	0.400	0.357	10.8	109	0.00
27 SURR	Fluoranthene-d10	0.400	0.373	6.8	110	0.00
28	Fluoranthene	0.400	0.368	8.0	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.373	6.8	108	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	109	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	101	0.00
33	Chrysene	0.400	0.391	2.3	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	102	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	103	0.00
36 I	Perylene-d12	0.400	0.400	0.0	98	0.00
37	Benzo(b)fluoranthene	0.400	0.374	6.5	108	0.00
38	Benzo(k)fluoranthene	0.400	0.387	3.3	108	0.00
39 C	Benzo(a)pyrene	0.400	0.384	4.0	104	0.00
40	Dibenzo(a,h)anthracene	0.400	0.383	4.3	105	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN121020