

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121020\
 Data File : BN013053.D
 Acq On : 10 Dec 2020 21:08
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 12/14/2020 2:42:56 PM

Quant Time: Dec 11 01:01:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 11 00:59:43 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.370	152	858	0.40	ng	0.00
7) Naphthalene-d8	10.137	136	2809	0.40	ng	# 0.00
13) Acenaphthene-d10	14.029	164	1754	0.40	ng	0.00
19) Phenanthrene-d10	16.800	188	3892	0.40	ng	# 0.00
29) Chrysene-d12	21.027	240	4359	0.40	ng	# 0.01
36) Perylene-d12	23.119	264	5071	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.994	112	272	0.09	ng	0.00
5) Phenol-d6	6.573	99	293	0.08	ng	0.00
8) Nitrobenzene-d5	8.540	82	205	0.06	ng	0.01
11) 2-Methylnaphthalene-d10	11.760	152	404	0.09	ng	0.02
14) 2,4,6-Tribromophenol	15.564	330	95	0.07	ng	0.01
15) 2-Fluorobiphenyl	12.659	172	615	0.09	ng	0.01
27) Fluoranthene-d10	18.846	212	1117	0.09	ng	0.00
31) Terphenyl-d14	19.466	244	907	0.09	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.965	88	278	0.21	ng	# 71
6) bis(2-Chloroethyl)ether	6.822	93	211	0.09	ng	92
9) Naphthalene	10.188	128	743	0.09	ng	# 70
10) Hexachlorobutadiene	10.454	225	157	0.09	ng	# 94
12) 2-Methylnaphthalene	11.831	142	493	0.10	ng	# 90
16) Acenaphthylene	13.750	152	753	0.09	ng	99
17) Acenaphthene	14.092	154	474	0.08	ng	86
18) Fluorene	15.107	166	639	0.09	ng	96
20) 4,6-Dinitro-2-methylph...	15.260	198	43	0.05	ng	# 1
21) 4-Bromophenyl-phenylether	16.021	248	182	0.08	ng	91
22) Hexachlorobenzene	16.106	284	244	0.08	ng	99
23) Atrazine	16.289	200	189	0.08	ng	# 70
25) Phenanthrene	16.836	178	1097	0.10	ng	97
26) Anthracene	16.934	178	853	0.08	ng	98
28) Fluoranthene	18.876	202	1359	0.10	ng	99
30) Pyrene	19.243	202	1476	0.10	ng	99
32) Benzo(a)anthracene	21.006	228	1323m	0.10	ng	
33) Chrysene	21.059	228	1376	0.09	ng	93
34) Bis(2-ethylhexyl)phtha...	20.953	149	1149	0.12	ng	# 88
35) Indeno(1,2,3-cd)pyrene	25.183	276	1787	0.08	ng	# 88
37) Benzo(b)fluoranthene	22.503	252	1558	0.10	ng	# 51
38) Benzo(k)fluoranthene	22.548	252	1605	0.09	ng	# 56
39) Benzo(a)pyrene	23.034	252	1472	0.09	ng	# 27
40) Dibenzo(a,h)anthracene	25.201	278	1555	0.09	ng	# 53
41) Benzo(g,h,i)perylene	25.800	276	1794	0.10	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121020\
Data File : BN013053.D
Acq On : 10 Dec 2020 21:08
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM

Quant Time: Dec 11 01:01:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 11 00:59:43 2020
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.370 min Scan# 570
 Delta R.T. -0.000 min
 Lab File: BN013053.D
 Acq: 10 Dec 2020 21:08




Tgt Ion: 152 Resp: 858
 Ion Ratio Lower Upper
 152 100
 150 147.8 116.6 174.8
 115 69.4 52.9 79.3




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED



mohammad
 12/14/2020 2:42:56 PM




#2
 1,4-Dioxane
 Concen: 0.21 ng
 RT: 2.965 min Scan# 23
 Delta R.T. 0.008 min
 Lab File: BN013053.D
 Acq: 10 Dec 2020 21:08



Tgt Ion: 88 Resp: 278
 Ion Ratio Lower Upper
 88 100
 58 38.5 59.4 89.2#
 43 43.2 32.3 48.5









#4
2-Fluorophenol
Concen: 0.09 ng
RT: 4.994 min Scan# 275
Delta R.T. 0.008 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 112 Resp: 272
Ion Ratio Lower Upper
112 100
64 65.8 49.5 74.3
63 40.8 32.0 48.0

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#5
Phenol-d6
Concen: 0.08 ng
RT: 6.573 min Scan# 471
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 99 Resp: 293
Ion Ratio Lower Upper
99 100
42 23.2 22.4 33.6
71 43.7 37.8 56.6









#6
bis(2-Chloroethyl)ether
Concen: 0.09 ng
RT: 6.822 min Scan# 502
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion:	93	Resp:	211
Ion	Ratio	Lower	Upper
93	100		
63	88.2	63.1	94.7
95	33.6	26.1	39.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.137 min Scan# 831
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08


Tgt Ion:	136	Resp:	2809
Ion	Ratio	Lower	Upper
136	100		
137	14.8	10.6	15.8
54	12.3	8.6	12.8
68	9.5	5.7	8.5





#8
Nitrobenzene-d5
Concen: 0.06 ng
RT: 8.540 min Scan# 705
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion: 82 Resp: 205
Ion Ratio Lower Upper
82 100
128 58.1 30.6 46.0#
54 66.1 48.3 72.5

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#9
Naphthalene
Concen: 0.09 ng
RT: 10.188 min Scan# 835
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion: 128 Resp: 743
Ion Ratio Lower Upper
128 100
129 25.2 9.8 14.8#
127 27.2 12.2 18.4#





#10
Hexachlorobutadiene
Concen: 0.09 ng
RT: 10.454 min Scan# 856
Delta R.T. -0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:225 Resp: 157
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 68.2 51.0 76.6

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#11
2-Methylnaphthalene-d10
Concen: 0.09 ng
RT: 11.760 min Scan# 996
Delta R.T. 0.018 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:152 Resp: 404
Ion Ratio Lower Upper
152 100
151 0.0 16.7 25.1#





#12
2-Methylnaphthalene
Concen: 0.10 ng
RT: 11.831 min Scan# 1012
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Tgt Ion:142 Resp: 493
Ion Ratio Lower Upper
142 100
141 90.1 69.4 104.2
115 59.9 35.7 53.5#

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.029 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08



Tgt Ion:164 Resp: 1754
Ion Ratio Lower Upper
164 100
162 106.0 80.6 120.8
160 53.3 39.5 59.3









#14
2,4,6-Tribromophenol
Concen: 0.07 ng
RT: 15.564 min Scan# 1405
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 330 Resp: 95
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.0 34.4 51.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM




#15
2-Fluorobiphenyl
Concen: 0.09 ng
RT: 12.659 min Scan# 1176
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 172 Resp: 615
Ion Ratio Lower Upper
172 100
171 48.9 28.2 42.4#
170 36.5 20.0 30.0#









#16
Acenaphthylene
Concen: 0.09 ng
RT: 13.750 min Scan# 1262
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08





Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.6	15.8	23.8
153	13.3	10.6	16.0



#17
Acenaphthene
Concen: 0.08 ng
RT: 14.092 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion	Ratio	Lower	Upper
154	100		
153	119.8	83.6	125.4
152	64.1	43.9	65.9



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
12/14/2020 2:42:56 PM



#18
Fluorene
Concen: 0.09 ng
RT: 15.107 min Scan# 1369
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:166 Resp: 639
Ion Ratio Lower Upper
166 100
165 102.5 78.6 118.0
167 14.2 10.7 16.1

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.800 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08



Tgt Ion:188 Resp: 3892
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.5 5.0 7.4#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.05 ng
RT: 15.260 min Scan# 1381
Delta R.T. 0.025 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion:198 Resp: 43
Ion Ratio Lower Upper
198 100
51 1752.7 43.5 65.3#
105 127.3 27.2 40.8#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#21
4-Bromophenyl-phenylether
Concen: 0.08 ng
RT: 16.021 min Scan# 1441
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08


Tgt Ion:248 Resp: 182
Ion Ratio Lower Upper
248 100
250 101.2 77.3 115.9
141 84.0 57.8 86.8





#22
Hexachlorobenzene
Concen: 0.08 ng
RT: 16.106 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Tgt Ion:284 Resp: 244
Ion Ratio Lower Upper
284 100
142 41.0 32.0 48.0
249 33.2 27.0 40.6

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#23
Atrazine
Concen: 0.08 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:200 Resp: 189
Ion Ratio Lower Upper
200 100
173 48.6 22.8 34.2#
215 64.2 38.1 57.1#









#25
Phenanthrene
Concen: 0.10 ng
RT: 16.836 min Scan# 1508
Delta R.T. 0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08





Tgt Ion:178 Resp: 1097
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 17.8 12.4 18.6



#26
Anthracene
Concen: 0.08 ng
RT: 16.934 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:178 Resp: 853
Ion Ratio Lower Upper
178 100
176 19.2 14.8 22.2
179 16.2 12.5 18.7



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
12/14/2020 2:42:56 PM



#27
Fluoranthene-d10
Concen: 0.09 ng
RT: 18.846 min Scan# 1772
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:212 Resp: 1117
Ion Ratio Lower Upper
212 100
106 9.3 6.8 10.2
104 6.7 4.1 6.1#

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#28
Fluoranthene
Concen: 0.10 ng
RT: 18.876 min Scan# 1778
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:202 Resp: 1359
Ion Ratio Lower Upper
202 100
101 8.0 5.8 8.6
203 17.1 13.7 20.5







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.027 min Scan# 2102
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:240 Resp: 4359
Ion Ratio Lower Upper
240 100
120 8.3 5.3 7.9#
236 27.7 21.8 32.8




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#30
Pyrene
Concen: 0.10 ng
RT: 19.243 min Scan# 1852
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:202 Resp: 1476
Ion Ratio Lower Upper
202 100
200 20.7 16.6 24.8
203 18.0 13.8 20.8






#31
Terphenyl-d14
Concen: 0.09 ng
RT: 19.466 min Scan# 1897
Delta R.T. 0.008 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08




Tgt Ion:244 Resp: 907
Ion Ratio Lower Upper
244 100
212 16.0 7.3 10.9#
122 16.9 7.3 10.9#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
12/14/2020 2:42:56 PM




#32
Benzo(a)anthracene
Concen: 0.10 ng m
RT: 21.006 min Scan# 2100
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08



Tgt Ion:228 Resp: 1323
Ion Ratio Lower Upper
228 100
226 31.4 22.3 33.5
229 24.7 15.7 23.5#









#33
Chrysene
Concen: 0.09 ng
RT: 21.059 min Scan# 2105
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion	Ratio	Lower	Upper
228	100		
226	33.1	24.2	36.2
229	23.5	15.7	23.5

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.12 ng
RT: 20.953 min Scan# 2095
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion	Ratio	Lower	Upper
149	100		
167	21.9	23.2	34.8#
279	3.8	3.9	5.9#









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.08 ng
RT: 25.183 min Scan# 3043
Delta R.T. 0.013 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 276 Resp: 1787
Ion Ratio Lower Upper
276 100
138 11.6 13.3 19.9#
277 19.2 20.2 30.4#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.119 min Scan# 2440
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion: 264 Resp: 5071
Ion Ratio Lower Upper
264 100
260 26.4 19.8 29.6
265 68.2 51.4 77.0









#37
Benzo(b)fluoranthene
Concen: 0.10 ng
RT: 22.503 min Scan# 2260
Delta R.T. 0.003 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion:	252	Resp:	1558
Ion Ratio	Lower	Upper	
252	100		
253	47.2	20.1	30.1#
125	32.2	7.2	10.8#

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
12/14/2020 2:42:56 PM



#38
Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 22.548 min Scan# 2273
Delta R.T. 0.006 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08

Tgt Ion:	252	Resp:	1605
Ion Ratio	Lower	Upper	
252	100		
253	47.1	19.8	29.8#
125	24.6	7.1	10.7#





#39

Benzo(a)pyrene

Concen: 0.09 ng

RT: 23.034 min Scan# 2415

Delta R.T. 0.006 min

Lab File: BN013053.D

Acq: 10 Dec 2020 21:08

Tgt Ion:	252	Resp:	1472
Ion Ratio	Lower	Upper	
252	100		
253	60.4	21.3	31.9#
125	44.7	8.3	12.5#

Instrument :





BNA_N

ClientSampleId :

SSTDIC0.1

Manual Integrations
APPROVED

mohammad
12/14/2020 2:42:56 PM



#40

Dibenzo(a,h)anthracene

Concen: 0.09 ng




RT: 25.201 min Scan# 3048


Delta R.T. 0.016 min

Lab File: BN013053.D


Acq: 10 Dec 2020 21:08

Tgt Ion:	278	Resp:	1555
Ion Ratio	Lower	Upper	
278	100		
139	24.3	9.3	13.9#
279	54.5	21.6	32.4#







#41
Benzo(g,h,i)perylene
Concen: 0.10 ng
RT: 25.800 min Scan# 3223
Delta R.T. 0.010 min
Lab File: BN013053.D
Acq: 10 Dec 2020 21:08



Tgt Ion: 276 Resp: 1794
Ion Ratio Lower Upper
276 100
277 33.4 19.8 29.8#
138 25.7 11.7 17.5#



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
12/14/2020 2:42:56 PM