

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121025\
 Data File : BN038307.D
 Acq On : 10 Dec 2025 15:18
 Operator : RC/JU
 Sample : Q3483-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Dec 10 15:42:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	6249	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	16861	0.400	ng	# 0.00
13) Acenaphthene-d10	14.398	164	8564	0.400	ng	0.00
19) Phenanthrene-d10	17.148	188	15660	0.400	ng	# 0.00
29) Chrysene-d12	21.348	240	13507	0.400	ng	# 0.00
35) Perylene-d12	23.645	264	14689	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	1430446	91.913	ng	0.00
5) Phenol-d6	6.894	99	1934545	104.563	ng	0.00
8) Nitrobenzene-d5	8.875	82	949778	66.782	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	26	0.001	ng	-0.02
14) 2,4,6-Tribromophenol	15.895	330	763944	201.068	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	2217649	53.751	ng	0.01
27) Fluoranthene-d10	19.006	212	3037	0.078	ng	-0.18
31) Terphenyl-d14	19.796	244	2273090	75.410	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.499	42	249	0.028	ng	# 1
6) bis(2-Chloroethyl)ether	7.147	93	2926200	161.682	ng	97
9) Naphthalene	10.583	128	5565377	111.661	ng	99
10) Hexachlorobutadiene	10.872	225	1184946	133.744	ng	# 100
12) 2-Methylnaphthalene	12.207	142	4392190	138.830	ng	97
16) Acenaphthylene	14.110	152	3505607	83.359	ng	99
17) Acenaphthene	14.462	154	3142123	107.240	ng	96
18) Fluorene	15.457	166	5610594	148.835	ng	95
21) 4-Bromophenyl-phenylether	16.466	248	26141	2.250	ng	# 1
22) Hexachlorobenzene	16.466	284	1528157	109.412	ng	99
24) Pentachlorophenol	16.801	266	1042316	191.240	ng	99
25) Phenanthrene	17.285	178	3974627	72.919	ng	99
26) Anthracene	17.285	178	3993563	84.982	ng	99
28) Fluoranthene	19.220	202	22706	0.392	ng	97
30) Pyrene	19.582	202	31769	0.478	ng	98
32) Benzo(a)anthracene	21.331	228	4841413	92.785	ng	98
33) Chrysene	21.384	228	2295529	39.956	ng	99
34) Bis(2-ethylhexyl)phtha...	21.268	149	46700	1.631	ng	98
36) Indeno(1,2,3-cd)pyrene	25.972	276	13162	0.169	ng	99
37) Benzo(b)fluoranthene	22.955	252	2958553	43.851	ng	# 89
38) Benzo(k)fluoranthene	23.005	252	5851760	86.998	ng	# 90
39) Benzo(a)pyrene	23.551	252	5195506	94.102	ng	# 81
40) Dibenzo(a,h)anthracene	25.981	278	3437	0.057	ng	# 38
41) Benzo(g,h,i)perylene	26.683	276	14705	0.219	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121025\
Data File : BN038307.D
Acq On : 10 Dec 2025 15:18
Operator : RC/JU
Sample : Q3483-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :


Quant Time: Dec 10 15:42:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.724 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:152 Resp: 6249
Ion Ratio Lower Upper
152 100
150 163.0 122.4 183.6
115 59.6 47.3 70.9




#3
n-Nitrosodimethylamine
Concen: 0.028 ng
RT: 3.499 min Scan# 58
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18



Tgt Ion: 42 Resp: 249
Ion Ratio Lower Upper
42 100
74 0.0 95.3 142.9#
44 0.0 7.9 11.9#









#4
2-Fluorophenol
Concen: 91.913 ng
RT: 5.290 min Scan# 306
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:112 Resp: 1430446
Ion Ratio Lower Upper
112 100
64 56.6 44.4 66.6
63 29.5 23.4 35.0



#5
Phenol-d6
Concen: 104.563 ng
RT: 6.894 min Scan# 528
Delta R.T. 0.007 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion: 99 Resp: 1934545
Ion Ratio Lower Upper
99 100
42 20.2 16.5 24.7
71 31.9 24.9 37.3









#6
bis(2-Chloroethyl)ether
Concen: 161.682 ng
RT: 7.147 min Scan# 563
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :

Tgt Ion: 93 Resp: 2926200




Ion	Ratio	Lower	Upper
93	100		
63	76.2	59.2	88.8
95	33.2	25.2	37.8




#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.530 min Scan# 932
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion: 136 Resp: 16861


Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	8.7	5.2	7.8#
68	6.7	4.1	6.1#








#8
Nitrobenzene-d5
Concen: 66.782 ng
RT: 8.875 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion: 82 Resp: 949778
Ion Ratio Lower Upper
82 100
128 41.3 34.0 51.0
54 56.7 44.4 66.6



#9
Naphthalene
Concen: 111.661 ng
RT: 10.583 min Scan# 937
Delta R.T. 0.011 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:128 Resp: 5565377
Ion Ratio Lower Upper
128 100
129 11.6 9.1 13.7
127 13.0 10.9 16.3








#10
Hexachlorobutadiene
Concen: 133.744 ng
RT: 10.872 min Scan# 90
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:225 Resp: 1184946
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.2 76.8



#11
2-Methylnaphthalene-d10
Concen: 0.001 ng
RT: 12.106 min Scan# 1141
Delta R.T. -0.020 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:152 Resp: 26
Ion Ratio Lower Upper
152 100
151 276.9 17.0 25.6#








#12
2-Methylnaphthalene
Concen: 138.830 ng
RT: 12.207 min Scan# 1161
Delta R.T. 0.005 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :



Tgt Ion:142 Resp: 4392190




Ion	Ratio	Lower	Upper
142	100		
141	86.6	71.4	107.2
115	33.7	28.4	42.6




#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.398 min Scan# 1449
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion:164 Resp: 8564


Ion	Ratio	Lower	Upper
164	100		
162	105.3	86.2	129.4
160	50.8	44.6	67.0








#14
2,4,6-Tribromophenol
Concen: 201.068 ng
RT: 15.895 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:330 Resp: 763944
Ion Ratio Lower Upper
330 100
332 96.8 75.8 113.6
141 37.7 31.8 47.8




#15
2-Fluorobiphenyl
Concen: 53.751 ng
RT: 13.019 min Scan# 1320
Delta R.T. 0.011 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Tgt Ion:172 Resp: 2217649
Ion Ratio Lower Upper
172 100
171 34.5 28.6 42.8
170 22.5 19.3 28.9








#16
Acenaphthylene
Concen: 83.359 ng
RT: 14.110 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:152 Resp: 3505607
Ion Ratio Lower Upper
152 100
151 19.9 15.4 23.2
153 13.5 10.2 15.4




#17
Acenaphthene
Concen: 107.240 ng
RT: 14.462 min Scan# 1455
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Tgt Ion:154 Resp: 3142123
Ion Ratio Lower Upper
154 100
153 110.7 89.8 134.8
152 53.2 47.5 71.3








#18
Fluorene
Concen: 148.835 ng
RT: 15.457 min Scan# 1548
Delta R.T. 0.011 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:166 Resp: 5610594
Ion Ratio Lower Upper
166 100
165 94.4 80.2 120.4
167 14.1 10.6 16.0



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.148 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion:188 Resp: 15660
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 15.1 7.4 11.0#









#21
4-Bromophenyl-phenylether
Concen: 2.250 ng
RT: 16.466 min Scan# 1630
Delta R.T. 0.124 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:248 Resp: 26141
Ion Ratio Lower Upper
248 100
250 152.8 78.2 117.2#
141 256.6 43.7 65.5#



#22
Hexachlorobenzene
Concen: 109.412 ng
RT: 16.466 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:284 Resp: 1528157
Ion Ratio Lower Upper
284 100
142 36.0 29.5 44.3
249 29.2 23.7 35.5








#24
Pentachlorophenol
Concen: 191.240 ng
RT: 16.801 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:266 Resp: 1042316
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.2
268 64.4 52.6 78.8




#25
Phenanthrene
Concen: 72.919 ng
RT: 17.285 min Scan# 1696
Delta R.T. 0.099 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Tgt Ion:178 Resp: 3974627
Ion Ratio Lower Upper
178 100
176 19.0 15.5 23.3
179 15.9 12.2 18.2








#26
 Anthracene
 Concen: 84.982 ng
 RT: 17.285 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BN038307.D
 Acq: 10 Dec 2025 15:18

Instrument :
 BNA_N
 ClientSampleId :



Tgt Ion:178 Resp: 3993563




Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.1	22.7
179	15.9	12.0	18.0




#27
 Fluoranthene-d10
 Concen: 0.078 ng
 RT: 19.006 min Scan# 1901
 Delta R.T. -0.181 min
 Lab File: BN038307.D
 Acq: 10 Dec 2025 15:18

Tgt Ion:212 Resp: 3037

Ion	Ratio	Lower	Upper
212	100		
106	0.0	13.7	20.5#
104	3.2	7.8	11.8#









#28
Fluoranthene
Concen: 0.392 ng
RT: 19.220 min Scan# 1946
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:202 Resp: 22706
Ion Ratio Lower Upper
202 100
101 14.1 9.6 14.4
203 17.1 13.5 20.3



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.348 min Scan# 2266
Delta R.T. 0.009 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion:240 Resp: 13507
Ion Ratio Lower Upper
240 100
120 25.2 15.4 23.2#
236 31.1 23.9 35.9









#30
Pyrene
Concen: 0.478 ng
RT: 19.582 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:202 Resp: 31769
Ion Ratio Lower Upper
202 100
200 21.4 16.9 25.3
203 19.7 14.2 21.2



#31
Terphenyl-d14
Concen: 75.410 ng
RT: 19.796 min Scan# 2071
Delta R.T. 0.005 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:244 Resp: 2273090
Ion Ratio Lower Upper
244 100
212 9.2 7.9 11.9
122 17.7 15.0 22.6








#32
Benzo(a)anthracene
Concen: 92.785 ng
RT: 21.331 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:228 Resp: 4841413
Ion Ratio Lower Upper
228 100
226 29.3 23.3 34.9
229 21.2 15.7 23.5



#33
Chrysene
Concen: 39.956 ng
RT: 21.384 min Scan# 2270
Delta R.T. 0.009 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion:228 Resp: 2295529
Ion Ratio Lower Upper
228 100
226 31.5 25.0 37.4
229 20.0 15.5 23.3









#34
Bis(2-ethylhexyl)phthalate
Concen: 1.631 ng
RT: 21.268 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:149 Resp: 46700
Ion Ratio Lower Upper
149 100
167 28.1 21.4 32.0
279 2.9 2.4 3.6



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.645 min Scan# 2808
Delta R.T. 0.012 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:264 Resp: 14689
Ion Ratio Lower Upper
264 100
260 28.9 22.2 33.2
265 85.2 80.2 120.2








#36
Indeno(1,2,3-cd)pyrene
Concen: 0.169 ng
RT: 25.972 min Scan# 30
Delta R.T. 0.009 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:276 Resp: 13162
Ion Ratio Lower Upper
276 100
138 29.1 23.6 35.4
277 24.4 19.7 29.5



#37
Benzo(b)fluoranthene
Concen: 43.851 ng
RT: 22.955 min Scan# 2572
Delta R.T. 0.009 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:252 Resp: 2958553
Ion Ratio Lower Upper
252 100
253 22.9 21.1 31.7
125 12.4 16.4 24.6#







#38
Benzo(k)fluoranthene
Concen: 86.998 ng
RT: 23.005 min Scan# 2589
Delta R.T. 0.015 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:252 Resp: 5851760
Ion Ratio Lower Upper
252 100
253 23.9 21.4 32.0
125 12.6 16.2 24.2#



#39
Benzo(a)pyrene
Concen: 94.102 ng
RT: 23.551 min Scan# 2776
Delta R.T. 0.018 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18


Tgt Ion:252 Resp: 5195506
Ion Ratio Lower Upper
252 100
253 23.6 23.3 34.9
125 13.5 22.4 33.6#








#40
Dibenzo(a,h)anthracene
Concen: 0.057 ng
RT: 25.981 min Scan# 3607
Delta R.T. -0.000 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Instrument :
BNA_N
ClientSampleId :



Tgt Ion:278 Resp: 3437
Ion Ratio Lower Upper
278 100
139 63.8 17.8 26.8#
279 48.1 21.3 31.9#



#41
Benzo(g,h,i)perylene
Concen: 0.219 ng
RT: 26.683 min Scan# 3847
Delta R.T. 0.006 min
Lab File: BN038307.D
Acq: 10 Dec 2025 15:18

Tgt Ion:276 Resp: 14705
Ion Ratio Lower Upper
276 100
277 26.7 20.0 30.0
138 33.5 21.8 32.6#

