




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121025\
Data File : BN038297.D
Acq On : 10 Dec 2025 09:12
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 10 10:07:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 10:07:34 2025
Response via : Initial Calibration



TIC: BN038297.D\data.ms

(70) Pentachlorophenol (C)




16.810min (0.000) 31792.54 ng

response 775196

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.41
264.00	61.60	63.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121025\
Data File : BN038297.D
Acq On : 10 Dec 2025 09:12
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 10 10:07:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 10:07:34 2025
Response via : Initial Calibration



TIC: BN038297.D\data.ms

(77) Benzidine

19.410min (0.000) 0.00 ng

response 3647174

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.89
183.00	13.20	12.19
0.00	0.00	0.00