

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121025\
 Data File : BN038300.D
 Acq On : 10 Dec 2025 11:05
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDICCC0.4

Quant Time: Dec 10 13:59:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.459	0.465	-1.3	100	0.00
3	n-Nitrosodimethylamine	0.575	0.568	1.2	99	0.00
4 S	2-Fluorophenol	0.996	0.988	0.8	100	0.00
5 S	Phenol-d6	1.184	1.151	2.8	100	0.00
6	bis(2-Chloroethyl)ether	1.158	1.166	-0.7	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.337	0.329	2.4	100	0.00
9	Naphthalene	1.182	1.168	1.2	100	0.00
10	Hexachlorobutadiene	0.210	0.210	0.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.564	0.549	2.7	100	0.00
12	2-Methylnaphthalene	0.751	0.731	2.7	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
14 S	2,4,6-Tribromophenol	0.177	0.167	5.6	100	0.00
15 S	2-Fluorobiphenyl	1.927	1.927	0.0	98	0.00
16	Acenaphthylene	1.964	1.903	3.1	100	0.00
17	Acenaphthene	1.369	1.342	2.0	100	-0.01
18	Fluorene	1.761	1.707	3.1	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.066	0.057	13.6	100	0.00
21	4-Bromophenyl-phenylether	0.297	0.287	3.4	100	0.00
22	Hexachlorobenzene	0.357	0.350	2.0	100	-0.01
23	Atrazine	0.202	0.190	5.9	100	0.00
24	Pentachlorophenol	0.139	0.121	12.9	100	0.00
25	Phenanthrene	1.392	1.363	2.1	100	0.00
26	Anthracene	1.200	1.146	4.5	100	-0.01
27 SURR	Fluoranthene-d10	0.989	0.984	0.5	100	0.00
28	Fluoranthene	1.480	1.459	1.4	100	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
30	Pyrene	1.969	1.920	2.5	100	0.00
31 S	Terphenyl-d14	0.893	0.881	1.3	100	0.00
32	Benzo(a)anthracene	1.545	1.479	4.3	100	0.00
33	Chrysene	1.701	1.721	-1.2	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.848	0.845	0.4	100	0.00
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	2.123	2.122	0.0	100	0.00
37	Benzo(b)fluoranthene	1.837	1.783	2.9	100	0.00
38	Benzo(k)fluoranthene	1.832	1.796	2.0	100	0.00
39 C	Benzo(a)pyrene	1.503	1.469	2.3	100	0.00
40	Dibenzo(a,h)anthracene	1.637	1.619	1.1	100	0.00
41	Benzo(g,h,i)perylene	1.829	1.812	0.9	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDICCC0.4