

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121118\
Quantitation Report (Not Reviewed)

Data File : BN003821.D
Acq On : 11 Dec 2018 18:58
Operator : JU/SJ
Sample : J6170-15
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9M10

Quant Time: Dec 12 02:56:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 02:52:45 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	4912	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	19176	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	9501	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	788482	0.40	ng/ul	0.09
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.71	264	14108	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	5502	0.19	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.56	252	2538	0.050	ng/ul#	Qvalue 27


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN003821.D
Acq On : 11 Dec 2018 18:58
Operator : JU/SJ
Sample : J6170-15
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9M10

Quant Time: Dec 12 02:56:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 02:52:45 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.050 ng/ul

RT: 23.56 min Scan# 5490

Delta R.T. -0.05 min

Lab File: BN003821.D


Acq: 11 Dec 2018 18:58

Instrument :

BNA_N

ClientSampleId :

F9M10




Tgt Ion: 252 Resp: 2538

Ion Ratio Lower Upper

252 100

253 52.3 19.0 28.4#

125 52.8 10.2 15.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

