

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121118\
Quantitation Report (Not Reviewed)

Data File : BN003833.D
Acq On : 12 Dec 2018 02:02
Operator : JU/SJ
Sample : J6171-14
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9Y18

Quant Time: Dec 12 03:49:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 03:46:49 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5035	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	19932	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	10186	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	899312	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.39
20) Perylene-d12	23.71	264	14608	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	6008	0.20	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.56	252	2675	0.051	ng/ul#	Qvalue 37


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN003833.D
Acq On : 12 Dec 2018 02:02
Operator : JU/SJ
Sample : J6171-14
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9Y18

Quant Time: Dec 12 03:49:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 03:46:49 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.051 ng/ul

RT: 23.56 min Scan# 5488

Delta R.T. -0.05 min

Lab File: BN003833.D


Acq: 12 Dec 2018 02:02

Instrument :

BNA_N

ClientSampleId :

F9Y18



Tgt Ion: 252 Resp: 2675

Ion Ratio Lower Upper

252 100

253 49.1 19.0 28.4#

125 46.1 10.2 15.4#

