

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121118\
 Data File : BN003824.D
 Acq On : 11 Dec 2018 20:44
 Operator : JU/SJ
 Sample : J6170-18
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9M13

Manual Integrations
 APPROVED

Sohil
 12/12/2018 5:50:26 PM

Quant Time: Dec 12 03:32:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	4973	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	18950	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	8969	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	16668m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	12480m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	12414m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	4832	0.17	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	7756m	0.16	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	17.25	178	1794m	0.033	ng/ul	
15) Fluoranthene	19.27	202	8896m	0.138	ng/ul	
17) Pyrene	19.63	202	7007m	0.157	ng/ul	
18) Benzo(a)anthracene	21.38	228	3780m	0.096	ng/ul	
19) Chrysene	21.43	228	5422m	0.120	ng/ul	
21) Benzo(b)fluoranthene	23.01	252	8683m	0.169	ng/ul	
22) Benzo(k)fluoranthene	23.06	252	3004m	0.060	ng/ul	
23) Benzo(a)pyrene	23.61	252	4796	0.108	ng/ul#	78
24) Indeno(1,2,3-cd)pyrene	26.06	276	4531	0.089	ng/ul#	95
25) Dibenzo(a,h)anthracene	26.07	278	1080	0.026	ng/ul#	31
26) Benzo(g,h,i)perylene	26.79	276	3925	0.089	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121118\
Data File : BN003824.D
Acq On : 11 Dec 2018 20:44
Operator : JU/SJ
Sample : J6170-18
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9M13

Quant Time: Dec 12 03:32:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 02:52:45 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/12/2018 5:50:26 PM

