

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121118\
 Data File : BN003831.D
 Acq On : 12 Dec 2018 00:52
 Operator : JU/SJ
 Sample : PB115306BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK06

Manual Integrations
 APPROVED

Sohil
 12/12/2018 5:50:45 PM

Quant Time: Dec 12 03:52:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 03:46:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	4661	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	18035	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	9279	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	17218m	0.40	ng/ul	0.00
16) Chrysene-d12	21.39	240	11180m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	10739m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.22	152	8078	0.29	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	13358m	0.27	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121118\
Data File : BN003831.D
Acq On : 12 Dec 2018 00:52
Operator : JU/SJ
Sample : PB115306BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK06

Manual Integrations
APPROVED

Sohil
12/12/2018 5:50:45 PM

Quant Time: Dec 12 03:52:09 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 03:46:49 2018
Response via : Initial Calibration

