

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
 Data File : BN038321.D
 Acq On : 11 Dec 2025 17:15
 Operator : RC/JU
 Sample : Q3787-13
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OW-03B-51.5-120425-FD

Quant Time: Dec 11 17:41:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5356	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	14557	0.400	ng	0.00
13) Acenaphthene-d10	14.398	164	7569	0.400	ng	0.00
19) Phenanthrene-d10	17.149	188	13759	0.400	ng	# 0.00
29) Chrysene-d12	21.349	240	11050	0.400	ng	# 0.00
35) Perylene-d12	23.639	264	11101	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	1956	0.147	ng	0.00
5) Phenol-d6	6.887	99	1433	0.090	ng	0.00
8) Nitrobenzene-d5	8.875	82	3982	0.324	ng	0.00
11) 2-Methylnaphthalene-d10	12.126	152	5840	0.284	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	1029	0.306	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	11508	0.316	ng	0.00
27) Fluoranthene-d10	19.187	212	12586	0.370	ng	0.00
31) Terphenyl-d14	19.787	244	10741	0.436	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.268	149	2724	0.116	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
Data File : BN038321.D
Acq On : 11 Dec 2025 17:15
Operator : RC/JU
Sample : Q3787-13
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OW-03B-51.5-120425-FD

Quant Time: Dec 11 17:41:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

