

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
 Data File : BN038325.D
 Acq On : 11 Dec 2025 19:40
 Operator : RC/JU
 Sample : Q3788-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OW-01B-66.5-120525

Quant Time: Dec 12 00:10:47 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5784	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	15998	0.400	ng	#-0.01
13) Acenaphthene-d10	14.388	164	8662	0.400	ng	-0.01
19) Phenanthrene-d10	17.149	188	16764	0.400	ng	0.00
29) Chrysene-d12	21.340	240	13742	0.400	ng	# 0.00
35) Perylene-d12	23.636	264	13571	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	2104	0.146	ng	0.00
5) Phenol-d6	6.887	99	1545	0.090	ng	0.00
8) Nitrobenzene-d5	8.875	82	4453	0.330	ng	0.00
11) 2-Methylnaphthalene-d10	12.126	152	6020	0.267	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	1270	0.330	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	11355	0.272	ng	0.00
27) Fluoranthene-d10	19.188	212	15855	0.382	ng	0.00
31) Terphenyl-d14	19.787	244	12212	0.398	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.196	88	506	0.076	ng	# 63
34) Bis(2-ethylhexyl)phtha...	21.268	149	2401	0.082	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
Data File : BN038325.D
Acq On : 11 Dec 2025 19:40
Operator : RC/JU
Sample : Q3788-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OW-01B-66.5-120525

Quant Time: Dec 12 00:10:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

