

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
 Data File : BN038332.D
 Acq On : 12 Dec 2025 01:13
 Operator : RC/JU
 Sample : Q3756-03DL 2X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BR-01-190-120225DL

Quant Time: Dec 12 04:16:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5762	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	15461	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	8101	0.400	ng	-0.01
19) Phenanthrene-d10	17.149	188	15232	0.400	ng	0.00
29) Chrysene-d12	21.340	240	11067	0.400	ng	0.00
35) Perylene-d12	23.630	264	11044	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	1060	0.074	ng	0.00
5) Phenol-d6	6.894	99	813	0.048	ng	0.00
8) Nitrobenzene-d5	8.875	82	1988	0.152	ng	0.00
11) 2-Methylnaphthalene-d10	12.126	152	2870	0.132	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	618	0.172	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	5793	0.148	ng	0.00
27) Fluoranthene-d10	19.187	212	6697	0.178	ng	0.00
31) Terphenyl-d14	19.787	244	5373	0.218	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.189	88	17328	2.623	ng	97
9) Naphthalene	10.573	128	1798	0.039	ng	# 88
12) 2-Methylnaphthalene	12.202	142	644	0.022	ng	# 83
34) Bis(2-ethylhexyl)phtha...	21.268	149	1375	0.059	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
Data File : BN038332.D
Acq On : 12 Dec 2025 01:13
Operator : RC/JU
Sample : Q3756-03DL 2X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BR-01-190-120225DL

Quant Time: Dec 12 04:16:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

