

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
 Data File : BN038333.D
 Acq On : 12 Dec 2025 01:49
 Operator : RC/JU
 Sample : Q3757-03DL 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BR-03D-490-120325DL

Quant Time: Dec 12 04:17:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	4723	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	12829	0.400	ng	0.00
13) Acenaphthene-d10	14.398	164	6799	0.400	ng	0.00
19) Phenanthrene-d10	17.148	188	13086	0.400	ng	# 0.00
29) Chrysene-d12	21.340	240	10141	0.400	ng	0.00
35) Perylene-d12	23.630	264	10159	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	286	0.024	ng	0.00
5) Phenol-d6	6.887	99	237	0.017	ng	0.00
8) Nitrobenzene-d5	8.875	82	384	0.035	ng	0.00
11) 2-Methylnaphthalene-d10	12.126	152	623	0.034	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	166	0.055	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	1335	0.041	ng	0.00
27) Fluoranthene-d10	19.183	212	1374	0.042	ng	0.00
31) Terphenyl-d14	19.787	244	1628	0.072	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.189	88	11881	2.194	ng	Qvalue 96
34) Bis(2-ethylhexyl)phtha...	21.268	149	1001	0.047	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121125\
Data File : BN038333.D
Acq On : 12 Dec 2025 01:49
Operator : RC/JU
Sample : Q3757-03DL 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BR-03D-490-120325DL

Quant Time: Dec 12 04:17:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

